

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031519\
 Data File : BE098983.D
 Acq On : 15 Mar 2019 22:59
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 18 07:07:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.789	0.882	-11.8	133	0.00
3	n-Nitrosodimethylamine	1.006	0.681	32.3#	76	0.04
4 S	2-Fluorophenol	1.196	0.913	23.7	84	0.02
5 S	Phenol-d6	1.689	1.454	13.9	98	0.01
6	bis(2-Chloroethyl)ether	1.304	1.069	18.0	97	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
8 S	Nitrobenzene-d5	0.419	0.349	16.7	119	0.01
9	Naphthalene	4.581	4.534	1.0	130	0.00
10	Hexachlorobutadiene	0.303	0.293	3.3	128	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.700	5.3	124	0.00
12	2-Methylnaphthalene	0.743	0.667	10.2	119	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	135	0.00
14 S	2,4,6-Tribromophenol	0.202	0.131	35.1#	88	0.00
15 S	2-Fluorobiphenyl	1.656	1.546	6.6	127	0.00
16	Acenaphthylene	2.057	1.866	9.3	128	0.00
17	Acenaphthene	1.220	1.161	4.8	129	-0.01
18	Fluorene	1.760	1.590	9.7	125	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.217	0.207	4.6	113	0.01
21	Hexachlorobenzene	0.258	0.249	3.5	118	0.00
22	Pentachlorophenol	0.052	0.022	57.7#	66	0.01
23	Phenanthrene	1.137	1.090	4.1	114	0.00
24	Anthracene	0.972	0.788	18.9	99	0.00
25 SURR	Fluoranthene-d10	6.385	5.551	13.1	104	0.00
26	Fluoranthene	1.605	1.388	13.5	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pvrene	1.224	1.271	-3.8	103	0.00
29 S	Terphenyl-d14	0.972	0.925	4.8	94	0.00
30	Benzo(a)anthracene	1.248	1.139	8.7	92	0.00
31	Chrysene	1.336	1.251	6.4	95	0.00
32	Bis(2-ethylhexyl)phthalate	2.657	2.028	23.7	82	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.211	14.1	86	0.00
34 I	Perylene-d12	1.000	1.000	0.0	99	0.00
35	Benzo(b)fluoranthene	1.315	1.258	4.3	98	0.00
36	Benzo(k)fluoranthene	1.377	1.425	-3.5	108	-0.01
37 C	Benzo(a)pyrene	1.261	1.260	0.1	103	0.00
38	Dibenzo(a,h)anthracene	1.193	0.896	24.9	76	0.00
39	Benzo(a,h,i)perylene	1.248	1.011	19.0	82	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031519\
Data File : BE098983.D
Acq On : 15 Mar 2019 22:59
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 18 07:07:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			