

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031519\
 Data File : BE098992.D
 Acq On : 16 Mar 2019 7:23
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 18 07:21:14 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	132	0.00
2	1,4-Dioxane	0.789	0.738	6.5	129	0.00
3	n-Nitrosodimethylamine	1.006	0.689	31.5#	89	0.04
4 S	2-Fluorophenol	1.196	0.967	19.1	103	0.01
5 S	Phenol-d6	1.689	1.552	8.1	122	0.00
6	bis(2-Chloroethyl)ether	1.304	0.991	24.0	104	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	145	0.00
8 S	Nitrobenzene-d5	0.419	0.356	15.0	137	0.00
9	Naphthalene	4.581	4.447	2.9	145	-0.01
10	Hexachlorobutadiene	0.303	0.282	6.9	140	-0.01
11 SURR	2-Methylnaphthalene-d10	0.739	0.696	5.8	140	0.00
12	2-Methylnaphthalene	0.743	0.707	4.8	143	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	147	0.00
14 S	2,4,6-Tribromophenol	0.202	0.142	29.7#	104	0.00
15 S	2-Fluorobiphenyl	1.656	1.590	4.0	143	-0.01
16	Acenaphthylene	2.057	1.960	4.7	147	-0.01
17	Acenaphthene	1.220	1.213	0.6	147	-0.01
18	Fluorene	1.760	1.623	7.8	139	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	137	0.00
20	4-Bromophenyl-phenylether	0.217	0.199	8.3	128	0.00
21	Hexachlorobenzene	0.258	0.244	5.4	136	0.00
22	Pentachlorophenol	0.052	0.022	57.7#	77	0.01
23	Phenanthrene	1.137	1.011	11.1	125	0.00
24	Anthracene	0.972	0.766	21.2	114	0.00
25 SURR	Fluoranthene-d10	6.385	5.312	16.8	118	0.00
26	Fluoranthene	1.605	1.329	17.2	118	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	121	-0.01
28	Pvrene	1.224	1.179	3.7	117	0.00
29 S	Terphenyl-d14	0.972	0.882	9.3	110	0.00
30	Benzo(a)anthracene	1.248	1.008	19.2	99	0.00
31	Chrysene	1.336	1.321	1.1	122	-0.01
32	Bis(2-ethylhexyl)phthalate	2.657	2.241	15.7	111	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.057	25.0#	92	0.00
34 I	Perylene-d12	1.000	1.000	0.0	122	-0.01
35	Benzo(b)fluoranthene	1.315	1.197	9.0	115	0.00
36	Benzo(k)fluoranthene	1.377	1.293	6.1	120	0.00
37 C	Benzo(a)pyrene	1.261	1.164	7.7	117	-0.01
38	Dibenzo(a,h)anthracene	1.193	0.880	26.2#	92	0.00
39	Benzo(a,h,i)perylene	1.248	1.005	19.5	101	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031519\
Data File : BE098992.D
Acq On : 16 Mar 2019 7:23
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 18 07:21:14 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		