

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031519\
 Data File : BE098992.D
 Acq On : 16 Mar 2019 7:23
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 18 07:21:14 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	132	0.00
2	1,4-Dioxane	0.400	0.375	6.3	129	0.00
3	n-Nitrosodimethylamine	0.400	0.274	31.5#	89	0.04
4 S	2-Fluorophenol	0.400	0.324	19.0	103	0.01
5 S	Phenol-d6	0.400	0.368	8.0	122	0.00
6	bis(2-Chloroethyl)ether	0.400	0.304	24.0	104	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	145	0.00
8 S	Nitrobenzene-d5	0.400	0.340	15.0	137	0.00
9	Naphthalene	0.400	0.388	3.0	145	-0.01
10	Hexachlorobutadiene	0.400	0.372	7.0	140	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	140	0.00
12	2-Methylnaphthalene	0.400	0.381	4.8	143	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	147	0.00
14 S	2,4,6-Tribromophenol	0.400	0.281	29.8#	104	0.00
15 S	2-Fluorobiphenyl	0.400	0.384	4.0	143	-0.01
16	Acenaphthylene	0.400	0.381	4.8	147	-0.01
17	Acenaphthene	0.400	0.398	0.5	147	-0.01
18	Fluorene	0.400	0.369	7.8	139	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	137	0.00
20	4-Bromophenyl-phenylether	0.400	0.367	8.3	128	0.00
21	Hexachlorobenzene	0.400	0.378	5.5	136	0.00
22	Pentachlorophenol	0.400	0.363	9.3	77	0.01
23	Phenanthrene	0.400	0.356	11.0	125	0.00
24	Anthracene	0.400	0.315	21.3	114	0.00
25 SURR	Fluoranthene-d10	0.400	0.333	16.8	118	0.00
26	Fluoranthene	0.400	0.331	17.3	118	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	121	-0.01
28	Pvrene	0.400	0.386	3.5	117	0.00
29 S	Terphenyl-d14	0.400	0.363	9.3	110	0.00
30	Benzo(a)anthracene	0.400	0.323	19.3	99	0.00
31	Chrysene	0.400	0.395	1.3	122	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.337	15.8	111	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.300	25.0#	92	0.00
34 I	Perylene-d12	0.400	0.400	0.0	122	-0.01
35	Benzo(b)fluoranthene	0.400	0.364	9.0	115	0.00
36	Benzo(k)fluoranthene	0.400	0.375	6.3	120	0.00
37 C	Benzo(a)pyrene	0.400	0.369	7.8	117	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.295	26.3#	92	0.00
39	Benzo(a,h,i)perylene	0.400	0.322	19.5	101	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031519\
Data File : BE098992.D
Acq On : 16 Mar 2019 7:23
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 18 07:21:14 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			