

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031819\
 Data File : BE099018.D
 Acq On : 19 Mar 2019 2:23
 Operator : JU/SJ
 Sample : K1937-19
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 BPS1-TT-MW2021-20190314

Quant Time: Mar 19 03:51:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.80	152	759	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	3611	0.40	ng	0.01
13) Acenaphthene-d10	14.46	164	2610	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	5898	0.40	ng	0.00
27) Chrysene-d12	21.40	240	6926	0.40	ng	0.00
34) Perylene-d12	23.90	264	6360	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	382	0.17	ng	0.03
5) Phenol-d6	7.01	99	264	0.08	ng	0.03
8) Nitrobenzene-d5	9.03	82	1152	0.30	ng	0.06
11) 2-Methylnaphthalene-d10	12.21	152	2495	0.37	ng	0.01
14) 2,4,6-Tribromophenol	15.97	330	467	0.35	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	4308	0.40	ng	0.00
25) Fluoranthene-d10	19.24	212	35255	0.37	ng	0.00
29) Terphenyl-d14	19.84	244	6446	0.38	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.28	88	526	0.346	ng	# 62
32) Bis(2-ethylhexyl)phthalate	21.29	149	3609	0.078	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE031819\
Data File : BE099018.D
Acq On : 19 Mar 2019 2:23
Operator : JU/SJ
Sample : K1937-19
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
BPS1-TT-MW2021-20190314

Quant Time: Mar 19 03:51:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

