

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE031819\
 Data File : BE098994.D
 Acq On : 18 Mar 2019 10:59
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 19 06:33:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.789	0.795	-0.8	110	0.00
3	n-Nitrosodimethylamine	1.006	0.401	60.1#	41#	0.05
4 S	2-Fluorophenol	1.196	0.919	23.2	78	0.04
5 S	Phenol-d6	1.689	1.517	10.2	94	0.01
6	bis(2-Chloroethyl)ether	1.304	1.031	20.9	86	0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
8 S	Nitrobenzene-d5	0.419	0.449	-7.2	137	0.01
9	Naphthalene	4.581	4.620	-0.9	119	0.00
10	Hexachlorobutadiene	0.303	0.298	1.7	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.739	0.725	1.9	116	0.00
12	2-Methylnaphthalene	0.743	0.702	5.5	113	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	122	0.00
14 S	2,4,6-Tribromophenol	0.202	0.129	36.1#	79	0.00
15 S	2-Fluorobiphenyl	1.656	2.284	-37.9#	171#	0.00
16	Acenaphthylene	2.057	2.010	2.3	125	0.00
17	Acenaphthene	1.220	1.211	0.7	122	-0.01
18	Fluorene	1.760	1.628	7.5	116	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4-Bromophenyl-phenylether	0.217	0.212	2.3	109	0.01
21	Hexachlorobenzene	0.258	0.259	-0.4	116	0.00
22	Pentachlorophenol	0.052	0.019	63.5#	54	0.04
23	Phenanthrene	1.137	1.074	5.5	106	0.00
24	Anthracene	0.972	0.802	17.5	95	0.00
25 SURR	Fluoranthene-d10	6.385	5.340	16.4	95	0.00
26	Fluoranthene	1.605	1.406	12.4	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
28	Pvrene	1.224	1.295	-5.8	97	0.00
29 S	Terphenyl-d14	0.972	2.990	-207.6#	282#	0.00
30	Benzo(a)anthracene	1.248	1.133	9.2	84	0.00
31	Chrysene	1.336	1.301	2.6	91	0.00
32	Bis(2-ethylhexyl)phthalate	2.657	1.802	32.2#	67	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.033	26.7#	68	0.01
34 I	Perylene-d12	1.000	1.000	0.0	91	0.00
35	Benzo(b)fluoranthene	1.315	1.254	4.6	90	0.00
36	Benzo(k)fluoranthene	1.377	1.530	-11.1	106	0.00
37 C	Benzo(a)pyrene	1.261	1.259	0.2	94	0.00
38	Dibenzo(a,h)anthracene	1.193	0.869	27.2#	67	0.00
39	Benzo(a,h,i)perylene	1.248	1.008	19.2	75	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE031819\
Data File : BE098994.D
Acq On : 18 Mar 2019 10:59
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 19 06:33:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			