

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE031819\
 Data File : BE099011.D
 Acq On : 18 Mar 2019 21:28
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 19 07:05:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	110	0.00
2	1,4-Dioxane	0.400	0.372	7.0	106	0.00
3	n-Nitrosodimethylamine	0.400	0.291	27.3#	79	0.03
4 S	2-Fluorophenol	0.400	0.292	27.0#	78	0.02
5 S	Phenol-d6	0.400	0.334	16.5	92	0.01
6	bis(2-Chloroethyl)ether	0.400	0.319	20.3	91	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	121	0.00
8 S	Nitrobenzene-d5	0.400	0.462	-15.5	156	0.01
9	Naphthalene	0.400	0.400	0.0	124	-0.01
10	Hexachlorobutadiene	0.400	0.384	4.0	120	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.382	4.5	118	0.00
12	2-Methylnaphthalene	0.400	0.365	8.8	114	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	124	0.00
14 S	2,4,6-Tribromophenol	0.400	0.320	20.0	100	0.00
15 S	2-Fluorobiphenyl	0.400	0.566	-41.5#	177	-0.01
16	Acenaphthylene	0.400	0.400	0.0	130	-0.01
17	Acenaphthene	0.400	0.404	-1.0	125	-0.01
18	Fluorene	0.400	0.385	3.8	122	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	0.00
20	4-Bromophenyl-phenylether	0.400	0.451	-12.7	125	0.00
21	Hexachlorobenzene	0.400	0.424	-6.0	122	0.00
22	Pentachlorophenol	0.400	0.399	0.3	84	0.00
23	Phenanthrene	0.400	0.403	-0.8	112	0.00
24	Anthracene	0.400	0.351	12.3	101	0.00
25 SURR	Fluoranthene-d10	0.400	0.360	10.0	101	0.00
26	Fluoranthene	0.400	0.377	5.8	106	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	99	-0.01
28	Pvrene	0.400	0.422	-5.5	105	0.00
29 S	Terphenyl-d14	0.400	1.198	-199.5#	300	0.00
30	Benzo(a)anthracene	0.400	0.334	16.5	85	0.00
31	Chrysene	0.400	0.436	-9.0	111	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.330	17.5	89	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.338	15.5	85	0.00
34 I	Perylene-d12	0.400	0.400	0.0	101	-0.01
35	Benzo(b)fluoranthene	0.400	0.406	-1.5	106	0.00
36	Benzo(k)fluoranthene	0.400	0.423	-5.7	112	-0.01
37 C	Benzo(a)pyrene	0.400	0.411	-2.7	108	0.00
38	Dibenzo(a,h)anthracene	0.400	0.284	29.0#	73	0.00
39	Benzo(a,h,i)perylene	0.400	0.351	12.3	91	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE031819\
Data File : BE099011.D
Acq On : 18 Mar 2019 21:28
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 19 07:05:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				