

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE031819\
 Data File : BE099011.D
 Acq On : 18 Mar 2019 21:28
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 19 07:05:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.789	0.734	7.0	106	0.00
3	n-Nitrosodimethylamine	1.006	0.732	27.2#	79	0.03
4 S	2-Fluorophenol	1.196	0.873	27.0#	78	0.02
5 S	Phenol-d6	1.689	1.412	16.4	92	0.01
6	bis(2-Chloroethyl)ether	1.304	1.039	20.3	91	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
8 S	Nitrobenzene-d5	0.419	0.485	-15.8	156#	0.01
9	Naphthalene	4.581	4.579	0.0	124	-0.01
10	Hexachlorobutadiene	0.303	0.291	4.0	120	-0.01
11 SURR	2-Methylnaphthalene-d10	0.739	0.707	4.3	118	0.00
12	2-Methylnaphthalene	0.743	0.678	8.7	114	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	124	0.00
14 S	2,4,6-Tribromophenol	0.202	0.162	19.8	100	0.00
15 S	2-Fluorobiphenyl	1.656	2.342	-41.4#	177#	-0.01
16	Acenaphthylene	2.057	2.059	-0.1	130	-0.01
17	Acenaphthene	1.220	1.231	-0.9	125	-0.01
18	Fluorene	1.760	1.693	3.8	122	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
20	4-Bromophenyl-phenylether	0.217	0.244	-12.4	125	0.00
21	Hexachlorobenzene	0.258	0.273	-5.8	122	0.00
22	Pentachlorophenol	0.052	0.030	42.3#	84	0.00
23	Phenanthrene	1.137	1.145	-0.7	112	0.00
24	Anthracene	0.972	0.853	12.2	101	0.00
25 SURR	Fluoranthene-d10	6.385	5.741	10.1	101	0.00
26	Fluoranthene	1.605	1.512	5.8	106	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	-0.01
28	Pvrene	1.224	1.292	-5.6	105	0.00
29 S	Terphenyl-d14	0.972	2.911	-199.5#	300#	0.00
30	Benzo(a)anthracene	1.248	1.044	16.3	85	0.00
31	Chrysene	1.336	1.456	-9.0	111	-0.01
32	Bis(2-ethylhexyl)phthalate	2.657	2.190	17.6	89	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.191	15.5	85	0.00
34 I	Perylene-d12	1.000	1.000	0.0	101	-0.01
35	Benzo(b)fluoranthene	1.315	1.335	-1.5	106	0.00
36	Benzo(k)fluoranthene	1.377	1.456	-5.7	112	-0.01
37 C	Benzo(a)pyrene	1.261	1.296	-2.8	108	0.00
38	Dibenzo(a,h)anthracene	1.193	0.846	29.1#	73	0.00
39	Benzo(a,h,i)perylene	1.248	1.095	12.3	91	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE031819\
Data File : BE099011.D
Acq On : 18 Mar 2019 21:28
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 19 07:05:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		