

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE031819\
 Data File : BE099013.D
 Acq On : 18 Mar 2019 23:18
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 19 02:49:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 15 05:31:33 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.789	0.730	7.5	109	0.00
3	n-Nitrosodimethylamine	1.006	0.459	54.4#	51	0.06
4 S	2-Fluorophenol	1.196	0.851	28.8#	78	0.02
5 S	Phenol-d6	1.689	1.492	11.7	100	0.01
6	bis(2-Chloroethyl)ether	1.304	1.032	20.9	93	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
8 S	Nitrobenzene-d5	0.419	0.452	-7.9	153#	0.01
9	Naphthalene	4.581	4.538	0.9	130	-0.01
10	Hexachlorobutadiene	0.303	0.276	8.9	120	-0.01
11 SURR	2-Methylnaphthalene-d10	0.739	0.689	6.8	121	0.00
12	2-Methylnaphthalene	0.743	0.691	7.0	122	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	128	0.00
14 S	2,4,6-Tribromophenol	0.202	0.144	28.7#	92	0.00
15 S	2-Fluorobiphenyl	1.656	2.362	-42.6#	186#	0.00
16	Acenaphthylene	2.057	2.005	2.5	131	0.00
17	Acenaphthene	1.220	1.209	0.9	128	-0.01
18	Fluorene	1.760	1.651	6.2	124	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.217	0.244	-12.4	126	0.00
21	Hexachlorobenzene	0.258	0.282	-9.3	127	0.00
22	Pentachlorophenol	0.052	0.024	53.8#	69	0.01
23	Phenanthrene	1.137	1.141	-0.4	113	0.00
24	Anthracene	0.972	0.836	14.0	100	0.00
25 SURR	Fluoranthene-d10	6.385	5.776	9.5	103	0.00
26	Fluoranthene	1.605	1.506	6.2	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	-0.01
28	Pvrene	1.224	1.323	-8.1	107	0.00
29 S	Terphenyl-d14	0.972	2.934	-201.9#	301#	0.00
30	Benzo(a)anthracene	1.248	1.156	7.4	94	0.00
31	Chrysene	1.336	1.296	3.0	98	-0.01
32	Bis(2-ethylhexyl)phthalate	2.657	2.004	24.6	81	-0.01
33	Indeno(1,2,3-cd)pyrene	1.410	1.178	16.5	84	0.00
34 I	Perylene-d12	1.000	1.000	0.0	99	-0.01
35	Benzo(b)fluoranthene	1.315	1.332	-1.3	104	0.00
36	Benzo(k)fluoranthene	1.377	1.539	-11.8	117	-0.01
37 C	Benzo(a)pyrene	1.261	1.329	-5.4	108	-0.01
38	Dibenzo(a,h)anthracene	1.193	0.765	35.9#	65	0.00
39	Benzo(a,h,i)perylene	1.248	1.119	10.3	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE031819\
Data File : BE099013.D
Acq On : 18 Mar 2019 23:18
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 19 02:49:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 15 05:31:33 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			