

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE031819\  
 Data File : BE099029.D  
 Acq On : 19 Mar 2019 9:06  
 Operator : JU/SJ  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 35 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Mar 20 06:57:55 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Mar 15 05:31:33 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	124	0.00
2	1,4-Dioxane	0.400	0.330	17.5	110	0.00
3	n-Nitrosodimethylamine	0.400	0.263	34.3#	80	0.05
4 S	2-Fluorophenol	0.400	0.263	34.3#	79	0.01
5 S	Phenol-d6	0.400	0.331	17.3	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.294	26.5#	95	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	132	0.00
8 S	Nitrobenzene-d5	0.400	0.428	-7.0	157	0.01
9	Naphthalene	0.400	0.402	-0.5	136	-0.01
10	Hexachlorobutadiene	0.400	0.377	5.8	128	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.387	3.3	130	0.00
12	2-Methylnaphthalene	0.400	0.375	6.3	127	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	137	0.00
14 S	2,4,6-Tribromophenol	0.400	0.299	25.3#	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.560	-40.0#	194	-0.01
16	Acenaphthylene	0.400	0.388	3.0	139	-0.01
17	Acenaphthene	0.400	0.395	1.3	136	-0.01
18	Fluorene	0.400	0.380	5.0	133	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	119	-0.02
20	4-Bromophenyl-phenylether	0.400	0.411	-2.7	124	0.00
21	Hexachlorobenzene	0.400	0.425	-6.2	133	-0.02
22	Pentachlorophenol	0.400	0.329	17.8	43	0.01
23	Phenanthrene	0.400	0.400	0.0	121	0.00
24	Anthracene	0.400	0.344	14.0	108	0.00
25 SURR	Fluoranthene-d10	0.400	0.357	10.8	110	0.00
26	Fluoranthene	0.400	0.370	7.5	114	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	104	-0.01
28	Pyrene	0.400	0.439	-9.7	114	0.00
29 S	Terphenyl-d14	0.400	1.209	-202.3#	316	-0.01
30	Benzo(a)anthracene	0.400	0.343	14.2	91	0.00
31	Chrysene	0.400	0.428	-7.0	114	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.312	22.0	88	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.344	14.0	91	0.00
34 I	Perylene-d12	0.400	0.400	0.0	108	-0.01
35	Benzo(b)fluoranthene	0.400	0.399	0.3	111	0.00
36	Benzo(k)fluoranthene	0.400	0.424	-6.0	120	-0.01
37 C	Benzo(a)pyrene	0.400	0.409	-2.2	114	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.291	27.3#	80	-0.01
39	Benzo(a,h,i)perylene	0.400	0.349	12.8	96	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE031819\  
Data File : BE099029.D  
Acq On : 19 Mar 2019 9:06  
Operator : JU/SJ  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 35 Sample Multiplier: 1

Instrument :  
BNA\_E  
LabSampleId :  
SSTDCCC0.4

Quant Time: Mar 20 06:57:55 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE030619.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Mar 15 05:31:33 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
-----					
(#) = Out of Range	SPCC's out = 0		CCC's out = 0		