

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032219\
 Data File : BE099088.D
 Acq On : 22 Mar 2019 20:08
 Operator : JU/SJ
 Sample : K1948-07
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 INV-025-SW01-031319

Quant Time: Mar 23 01:15:12 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 23 00:46:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	3177	0.40	ng	0.00
7) Naphthalene-d8	10.64	136	12019	0.40	ng	0.00
13) Acenaphthene-d10	14.48	164	7510	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	20783	0.40	ng	0.00
27) Chrysene-d12	21.39	240	30639	0.40	ng	0.00
34) Perylene-d12	23.90	264	35178	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.42	112	3929	0.42	ng	0.00
5) Phenol-d6	7.00	99	5649	0.41	ng	0.00
8) Nitrobenzene-d5	8.99	82	2980	0.57	ng	0.00
11) 2-Methylnaphthalene-d10	12.22	152	8167	0.38	ng	-0.02
14) 2,4,6-Tribromophenol	15.96	330	1064	0.48	ng	0.00
15) 2-Fluorobiphenyl	13.12	172	11366	0.38	ng	0.00
25) Fluoranthene-d10	19.24	212	102896	0.37	ng	0.00
29) Terphenyl-d14	19.84	244	17043	0.26	ng	0.00


Target Compounds				Qvalue		
2) 1,4-Dioxane	3.32	88	142	0.029	ng	# 26
3) n-Nitrosodimethylamine	3.67	42	164	0.064	ng	# 1
6) bis(2-Chloroethyl)ether	7.27	93	40524	3.661	ng	# 1
16) Acenaphthylene	14.21	152	1266	0.032	ng	# 92
22) Pentachlorophenol	16.87	266	71	0.267	ng	90
23) Phenanthrene	17.25	178	10856	0.176	ng	97
24) Anthracene	17.35	178	1241	0.021	ng	# 94
26) Fluoranthene	19.27	202	22283	0.257	ng	98
28) Pyrene	19.63	202	24130	0.232	ng	95
30) Benzo(a)anthracene	21.37	228	14663	0.136	ng	98
31) Chrysene	21.43	228	16774	0.156	ng	98
32) Bis(2-ethylhexyl)phthalate	21.31	149	23926	0.171	ng	99
33) Indeno(1,2,3-cd)pyrene	26.56	276	10051	0.087	ng	99
35) Benzo(b)fluoranthene	23.12	252	20978	0.172	ng	# 94
36) Benzo(k)fluoranthene	23.12	252	20978	0.176	ng	# 96
37) Benzo(a)pyrene	23.79	252	14202	0.126	ng	95
38) Dibenzo(a,h)anthracene	26.57	278	2633	0.024	ng	# 67
39) Benzo(g,h,i)perylene	27.38	276	10694	0.095	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032219\
Data File : BE099088.D
Acq On : 22 Mar 2019 20:08
Operator : JU/SJ
Sample : K1948-07
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
INV-025-SW01-031319

Quant Time: Mar 23 01:15:12 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 23 00:46:37 2019
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 0.400 ng

RT: 7.84 min Scan# 432

Delta R.T. -0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08

Instrument :

BNA_E

ClientSampleId :

INV-025-SW01-031319


Tgt Ion:152 Resp: 3177

Ion Ratio Lower Upper

152 100

150 144.9 122.5 183.7




115 61.6 57.4 86.2



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.029 ng

RT: 3.32 min Scan# 40

Delta R.T. -0.02 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08


Tgt Ion: 88 Resp: 142

Ion Ratio Lower Upper

88 100

58 0.0 52.0 78.0#



43 0.0 24.0 36.0#






Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#3

n-Nitrosodimethylamine

Concen: 0.064 ng

RT: 3.67 min Scan# 70

Delta R.T. -0.01 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08

Instrument :

BNA_E

ClientSampleId :

INV-025-SW01-031319





Tgt Ion: 42 Resp: 164

Ion Ratio Lower Upper

42 100

74 0.0 160.0 240.0#

44 0.0 0.0 0.0



#4

2-Fluorophenol

Concen: 0.415 ng

RT: 5.42 min Scan# 222

Delta R.T. 0.01 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08


Tgt Ion: 112 Resp: 3929


Ion Ratio Lower Upper

112 100

64 58.7 57.9 86.9

63 120.9 158.3 237.5#





#5

Phenol-d6

Concen: 0.415 ng

RT: 7.00 min Scan# 359

Delta R.T. 0.01 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08

Instrument :

BNA_E

ClientSampleId :

INV-025-SW01-031319


Tgt Ion: 99 Resp: 5649

Ion Ratio Lower Upper

99 100

42 36.5 24.2 36.2#




71 0.0 35.7 53.5#



Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 3.661 ng

RT: 7.27 min Scan# 383

Delta R.T. -0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08


Tgt Ion: 93 Resp: 40524

Ion Ratio Lower Upper

93 100

63 7.2 247.5 371.3#



95 2.2 21.8 32.8#




Abundance Ion 93.00 (92.70 to 93.70): BE0

Ion 63.00 (62.70 to 63.70): BE0

Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.64 min Scan# 652

Delta R.T. -0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08

Instrument :

BNA_E

ClientSampleId :

INV-025-SW01-031319

Tgt Ion:136 Resp: 12019





Ion Ratio Lower Upper

136 100

137 11.7 11.4 17.0

54 20.2 39.3 58.9#

68 5.7 6.7 10.1#



#8

Nitrobenzene-d5

Concen: 0.567 ng

RT: 8.99 min Scan# 525

Delta R.T. -0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08




Tgt Ion: 82 Resp: 2980


Ion Ratio Lower Upper

82 100

128 112.7 123.8 185.8#

54 172.2 203.4 305.2#





#11

2-Methylnaphthalene-d10

Concen: 0.377 ng

RT: 12.22 min Scan# 764

Delta R.T. -0.02 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08

Instrument :

BNA_E

ClientSampleId :


INV-025-SW01-031319

Tgt Ion:152 Resp: 8167

Ion Ratio Lower Upper

152 100

151 19.3 16.5 24.7



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

12.22

4000

3000



2000

1000

0

12.20 12.22 12.30

Time-->



#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.48 min Scan# 921

Delta R.T. -0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08


Tgt Ion:164 Resp: 7510

Ion Ratio Lower Upper

164 100

162 98.7 80.2 120.2

160 46.1 37.8 56.8



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E

14.48

6000


4000


2000

0

14.30 14.40 14.50 14.60

Time-->









#14
 2,4,6-Tribromophenol
 Concen: 0.482 ng
 RT: 15.96 min Scan# 1025
 Delta R.T. 0.00 min
 Lab File: BE099088.D
 Acq: 22 Mar 2019 20:08




Instrument :
 BNA_E
 ClientSampleId :
 INV-025-SW01-031319


Tgt Ion: 330 Resp: 1064
 Ion Ratio Lower Upper
 330 100
 332 94.9 77.6 116.4
 141 38.5 31.0 46.4



#15
 2-Fluorobiphenyl
 Concen: 0.381 ng
 RT: 13.12 min Scan# 826
 Delta R.T. -0.00 min
 Lab File: BE099088.D
 Acq: 22 Mar 2019 20:08

Tgt Ion: 172 Resp: 11366
 Ion Ratio Lower Upper
 172 100
 171 34.0 28.0 42.0
 170 21.9 17.7 26.5





#16

Acenaphthylene

Concen: 0.032 ng

RT: 14.21 min Scan# 902

Delta R.T. -0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08

Instrument :

BNA_E

ClientSampleId :

INV-025-SW01-031319


Tgt Ion:152 Resp: 1266

Ion Ratio Lower Upper

152 100

151 16.7 17.0 25.6#

153 16.4 11.2 16.8



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E

800

600

400



200

0

14.10 14.20 14.30

14.21

Time-->



#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.21 min Scan# 1119

Delta R.T. 0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08


Tgt Ion:188 Resp: 20783

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 8.6 6.3 9.5



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0

20000

15000

10000


5000


0

17.10 17.20 17.30

17.21

Time-->





#22

Pentachlorophenol

Concen: 0.267 ng

RT: 16.87 min Scan# 1093

Delta R.T. 0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08


Instrument :

BNA_E




ClientSampleId :

INV-025-SW01-031319

Tgt Ion:	266	Resp:	71
Ion	Ratio	Lower	Upper
266	100		
264	74.7	66.6	100.0
268	78.5	69.8	104.8



Abundance Ion 266.00 (265.70 to 266.70): E
Ion 264.00 (263.70 to 264.70): E
Ion 268.00 (267.70 to 268.70): E



#23

Phenanthrene

Concen: 0.176 ng


RT: 17.25 min Scan# 1122

Delta R.T. 0.00 min



Lab File: BE099088.D


Acq: 22 Mar 2019 20:08

Tgt Ion:	178	Resp:	10856
Ion	Ratio	Lower	Upper
178	100		
176	18.7	16.0	24.0
179	16.7	12.1	18.1



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E





#24

Anthracene

Concen: 0.021 ng

RT: 17.35 min Scan# 1129

Delta R.T. 0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08

Instrument :

BNA_E

ClientSampleId :

INV-025-SW01-031319


Tgt Ion:178 Resp: 1241

Ion Ratio Lower Upper

178 100

176 18.1 15.4 23.0




179 19.9 12.2 18.4#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#25

Fluoranthene-d10

Concen: 0.371 ng

RT: 19.24 min Scan# 1388

Delta R.T. 0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08


Tgt Ion:212 Resp: 102896

Ion Ratio Lower Upper

212 100

106 2.7 1.8 2.8



104 1.6 1.0 1.6#




Abundance Ion 212.00 (211.70 to 212.70): E

Ion 106.00 (105.70 to 106.70): E

Ion 104.00 (103.70 to 104.70): E









#26
 Fluoranthene
 Concen: 0.257 ng
 RT: 19.27 min Scan# 1393
 Delta R.T. -0.00 min
 Lab File: BE099088.D
 Acq: 22 Mar 2019 20:08




Instrument :
 BNA_E
 ClientSampleId :
 INV-025-SW01-031319


Tgt Ion: 202 Resp: 22283
 Ion Ratio Lower Upper
 202 100
 101 10.3 7.3 10.9
 203 17.7 13.4 20.2



#27
 Chrysene-d12
 Concen: 0.400 ng
 RT: 21.39 min Scan# 1632
 Delta R.T. 0.00 min
 Lab File: BE099088.D
 Acq: 22 Mar 2019 20:08

Tgt Ion: 240 Resp: 30639
 Ion Ratio Lower Upper
 240 100
 120 4.6 5.0 7.4#
 236 27.5 22.7 34.1





#28

Pyrene

Concen: 0.232 ng

RT: 19.63 min Scan# 1456

Delta R.T. 0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08





Instrument :

BNA_E

ClientSampleId :

INV-025-SW01-031319

Tgt Ion:	202	Resp:	24130
Ion	Ratio	Lower	Upper
202	100		
200	20.1	17.0	25.6
203	20.9	14.2	21.2



#29

Terphenyl-d14

Concen: 0.263 ng




RT: 19.84 min Scan# 1492


Delta R.T. 0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08

Tgt Ion:	244	Resp:	17043
Ion	Ratio	Lower	Upper
244	100		
212	30.4	29.4	44.2
122	9.0	7.1	10.7





#30

Benzo(a)anthracene

Concen: 0.136 ng

RT: 21.37 min Scan# 1630

Delta R.T. 0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08

Instrument :

BNA_E

ClientSampleId :

INV-025-SW01-031319


Tgt Ion:228 Resp: 14663

Ion Ratio Lower Upper

228 100

226 29.4 23.0 34.4




229 21.6 16.5 24.7



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#31

Chrysene

Concen: 0.156 ng

RT: 21.43 min Scan# 1635

Delta R.T. 0.00 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08


Tgt Ion:228 Resp: 16774

Ion Ratio Lower Upper

228 100

226 30.0 23.7 35.5



229 22.1 16.1 24.1




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#32
 Bis(2-ethylhexyl)phthalate
 Concen: 0.171 ng
 RT: 21.31 min Scan# 1625
 Delta R.T. 0.00 min
 Lab File: BE099088.D
 Acq: 22 Mar 2019 20:08

Instrument :
 BNA_E
 ClientSampleId :
 INV-025-SW01-031319

Tgt Ion:149 Resp: 23926
 Ion Ratio Lower Upper
 149 100
 167 7.2 5.4 8.0
 279 1.0 1.0 1.4




Abundance Ion 149.00 (148.70 to 149.70): E
 Ion 167.00 (166.70 to 167.70): E
 Ion 279.00 (278.70 to 279.70): E






#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.087 ng
 RT: 26.56 min Scan# 2950
 Delta R.T. -0.00 min
 Lab File: BE099088.D
 Acq: 22 Mar 2019 20:08

Tgt Ion:276 Resp: 10051
 Ion Ratio Lower Upper
 276 100
 138 17.8 13.8 20.8
 277 23.8 19.5 29.3



Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 138.00 (137.70 to 138.70): E
 Ion 277.00 (276.70 to 277.70): E









#34
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.90 min Scan# 2209
 Delta R.T. -0.00 min
 Lab File: BE099088.D
 Acq: 22 Mar 2019 20:08




Instrument :
 BNA_E
 ClientSampleId :
 INV-025-SW01-031319


Tgt Ion: 264 Resp: 35178
 Ion Ratio Lower Upper
 264 100
 260 24.3 21.2 31.8
 265 26.1 24.6 36.8



#35
 Benzo(b)fluoranthene
 Concen: 0.172 ng
 RT: 23.12 min Scan# 1991
 Delta R.T. -0.00 min
 Lab File: BE099088.D
 Acq: 22 Mar 2019 20:08

Tgt Ion: 252 Resp: 20978
 Ion Ratio Lower Upper
 252 100
 253 25.1 18.2 27.2
 125 12.2 7.3 10.9#









#36
Benzo(k)fluoranthene
Concen: 0.176 ng
RT: 23.12 min Scan# 1991
Delta R.T. -0.05 min
Lab File: BE099088.D
Acq: 22 Mar 2019 20:08




Instrument :
BNA_E
ClientSampleId :
INV-025-SW01-031319


Tgt Ion: 252 Resp: 20978
Ion Ratio Lower Upper
252 100
253 25.1 19.3 28.9
125 12.2 7.5 11.3#



#37
Benzo(a)pyrene
Concen: 0.126 ng
RT: 23.79 min Scan# 2177
Delta R.T. -0.00 min
Lab File: BE099088.D
Acq: 22 Mar 2019 20:08

Tgt Ion: 252 Resp: 14202
Ion Ratio Lower Upper
252 100
253 27.2 19.9 29.9
125 13.0 9.0 13.4





#38

Dibenzo(a,h)anthracene

Concen: 0.024 ng

RT: 26.57 min Scan# 2953

Delta R.T. -0.03 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08

Instrument :

BNA_E

ClientSampleId :

INV-025-SW01-031319


Tgt Ion: 278 Resp: 2633

Ion Ratio Lower Upper




278 100

139 34.5 11.8 17.6#

279 39.2 21.1 31.7#



Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#39

Benzo(g,h,i)perylene

Concen: 0.095 ng

RT: 27.38 min Scan# 3179

Delta R.T. -0.01 min

Lab File: BE099088.D

Acq: 22 Mar 2019 20:08


Tgt Ion: 276 Resp: 10694

Ion Ratio Lower Upper

276 100

277 26.1 20.4 30.6

138 18.9 13.6 20.4



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

