

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032219\
 Data File : BE099087.D
 Acq On : 22 Mar 2019 19:33
 Operator : JU/SJ
 Sample : K1948-06
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PST-028-SW064-031219

Manual Integrations
 APPROVED

Sohil
 3/25/2019 12:11:53 PM

Quant Time: Mar 23 01:25:00 2019

Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Sat Mar 23 00:46:37 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	2973	0.40	ng	0.00
7) Naphthalene-d8	10.64	136	11289	0.40	ng	0.00
13) Acenaphthene-d10	14.49	164	7155	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	20934	0.40	ng	0.00
27) Chrysene-d12	21.39	240	31630	0.40	ng	0.00
34) Perylene-d12	23.90	264	34707	0.40	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.41	112	3983	0.45	ng	0.00
5) Phenol-d6	6.99	99	5822	0.46	ng	0.00
8) Nitrobenzene-d5	8.99	82	2950m	0.60	ng	0.00
11) 2-Methylnaphthalene-d10	12.22	152	8436	0.41	ng	-0.01
14) 2,4,6-Tribromophenol	15.96	330	1117	0.53	ng	0.00
15) 2-Fluorobiphenyl	13.12	172	11879	0.42	ng	0.00
25) Fluoranthene-d10	19.24	212	124475	0.45	ng	0.00
29) Terphenyl-d14	19.84	244	21198	0.32	ng	0.00

Target Compounds

						Qvalue
23) Phenanthrene	17.25	178	4861	0.078	ng	99
26) Fluoranthene	19.27	202	10707	0.122	ng	98
28) Pyrene	19.63	202	9684	0.090	ng	96
30) Benzo(a)anthracene	21.37	228	6091	0.055	ng	98
31) Chrysene	21.43	228	7047m	0.063	ng	
32) Bis(2-ethylhexyl)phthalate	21.31	149	18700	0.129	ng	98
33) Indeno(1,2,3-cd)pyrene	26.55	276	3865	0.032	ng	# 84
35) Benzo(b)fluoranthene	23.13	252	8577	0.071	ng	# 83
36) Benzo(k)fluoranthene	23.17	252	3709m	0.032	ng	
37) Benzo(a)pyrene	23.79	252	5616	0.051	ng	# 86
39) Benzo(g,h,i)perylene	27.38	276	4084	0.037	ng	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032219\
Data File : BE099087.D
Acq On : 22 Mar 2019 19:33
Operator : JU/SJ
Sample : K1948-06
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PST-028-SW064-031219

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:53 PM

Quant Time: Mar 23 01:25:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 23 00:46:37 2019
Response via : Initial Calibration

