

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032219\
 Data File : BE099097.D
 Acq On : 23 Mar 2019 1:49
 Operator : JU/SJ
 Sample : K1947-06
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PST-025-B035-031319

Quant Time: Mar 23 04:44:33 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 23 04:40:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	3510	0.40	ng	0.00
7) Naphthalene-d8	10.64	136	13631	0.40	ng	0.00
13) Acenaphthene-d10	14.49	164	8915	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	23866	0.40	ng	0.00
27) Chrysene-d12	21.39	240	35939	0.40	ng	0.00
34) Perylene-d12	23.90	264	40748	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.41	112	4806	0.46	ng	0.00
5) Phenol-d6	6.99	99	6927	0.46	ng	0.00
8) Nitrobenzene-d5	8.99	82	4177	0.70	ng	-0.01
11) 2-Methylnaphthalene-d10	12.22	152	9168	0.37	ng	-0.01
14) 2,4,6-Tribromophenol	15.96	330	1901	0.73	ng	0.00
15) 2-Fluorobiphenyl	13.12	172	14826	0.42	ng	0.00
25) Fluoranthene-d10	19.24	212	116735	0.37	ng	0.00
29) Terphenyl-d14	19.84	244	28005	0.37	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
23) Phenanthrene	17.25	178	3341	0.047	ng	# 93
26) Fluoranthene	19.27	202	5922	0.059	ng	# 88
28) Pyrene	19.63	202	5696	0.047	ng	# 95
30) Benzo(a)anthracene	21.37	228	2940	0.023	ng	# 87
31) Chrysene	21.43	228	3746	0.030	ng	# 92
32) Bis(2-ethylhexyl)phthalate	21.31	149	23339	0.142	ng	# 98
37) Benzo(a)pyrene	23.78	252	3079	0.024	ng	# 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032219\
Data File : BE099097.D
Acq On : 23 Mar 2019 1:49
Operator : JU/SJ
Sample : K1947-06
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PST-025-B035-031319

Quant Time: Mar 23 04:44:33 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 23 04:40:11 2019
Response via : Initial Calibration

