

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032219\
 Data File : BE099082.D
 Acq On : 22 Mar 2019 16:10
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE032219

Quant Time: Mar 23 00:49:46 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 23 00:46:37 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.621	0.519	16.4	94	0.00
3	n-Nitrosodimethylamine	0.320	0.202	36.9#	67	-0.01
4 S	2-Fluorophenol	1.191	0.995	16.5	91	0.00
5 S	Phenol-d6	1.715	1.436	16.3	93	0.00
6	bis(2-Chloroethyl)ether	1.394	1.232	11.6	96	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.175	0.183	-4.6	124	0.00
9	Naphthalene	4.714	4.158	11.8	94	0.00
10	Hexachlorobutadiene	0.231	0.207	10.4	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.720	0.644	10.6	96	-0.01
12	2-Methylnaphthalene	0.807	0.687	14.9	93	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
14 S	2,4,6-Tribromophenol	0.118	0.106	10.2	111	0.00
15 S	2-Fluorobiphenyl	1.591	1.431	10.1	94	0.00
16	Acenaphthylene	2.135	1.770	17.1	94	0.00
17	Acenaphthene	1.266	1.089	14.0	96	0.00
18	Fluorene	1.847	1.590	13.9	96	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4-Bromophenyl-phenylether	0.221	0.189	14.5	96	0.00
21	Hexachlorobenzene	0.206	0.177	14.1	98	0.00
22	Pentachlorophenol	0.061	0.047	23.0	108	0.00
23	Phenanthrene	1.188	1.033	13.0	99	0.00
24	Anthracene	1.124	0.929	17.3	95	0.00
25 SURR	Fluoranthene-d10	5.336	4.699	11.9	97	0.00
26	Fluoranthene	1.671	1.419	15.1	98	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pyrene	1.357	1.121	17.4	98	0.00
29 S	Terphenyl-d14	0.846	0.740	12.5	99	0.00
30	Benzo(a)anthracene	1.409	1.187	15.8	99	0.00
31	Chrysene	1.408	1.220	13.4	103	0.00
32	Bis(2-ethylhexyl)phthalate	1.829	1.233	32.6#	95	0.00
33	Indeno(1,2,3-cd)pyrene	1.514	1.299	14.2	102	0.00
34 I	Perylene-d12	1.000	1.000	0.0	105	0.00
35	Benzo(b)fluoranthene	1.384	1.137	17.8	104	0.00
36	Benzo(k)fluoranthene	1.356	1.122	17.3	104	0.00
37 C	Benzo(a)pyrene	1.279	1.029	19.5	103	0.00
38	Dibenzo(a,h)anthracene	1.244	1.015	18.4	104	-0.01
39	Benzo(g,h,i)perylene	1.278	1.053	17.6	104	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032219\
Data File : BE099082.D
Acq On : 22 Mar 2019 16:10
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE032219

Quant Time: Mar 23 00:49:46 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 23 00:46:37 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			