

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032219\
 Data File : BE099101.D
 Acq On : 23 Mar 2019 4:18
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 07:08:46 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 23 04:40:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.621	0.510	17.9	109	0.00
3	n-Nitrosodimethylamine	0.320	0.237	25.9#	92	-0.01
4 S	2-Fluorophenol	1.191	1.003	15.8	109	0.00
5 S	Phenol-d6	1.715	1.479	13.8	113	0.00
6	bis(2-Chloroethyl)ether	1.394	1.225	12.1	113	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
8 S	Nitrobenzene-d5	0.175	0.231	-32.0#	206#	-0.01
9	Naphthalene	4.714	3.747	20.5	111	0.00
10	Hexachlorobutadiene	0.231	0.185	19.9	111	0.00
11 SURR	2-Methylnaphthalene-d10	0.720	0.595	17.4	117	-0.02
12	2-Methylnaphthalene	0.807	0.681	15.6	121	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	147	0.00
14 S	2,4,6-Tribromophenol	0.118	0.145	-22.9	234#	0.00
15 S	2-Fluorobiphenyl	1.591	1.187	25.4#	120	0.00
16	Acenaphthylene	2.135	1.610	24.6	131	-0.02
17	Acenaphthene	1.266	0.953	24.7	130	0.00
18	Fluorene	1.847	1.417	23.3	131	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	147	0.00
20	4-Bromophenyl-phenylether	0.221	0.178	19.5	139	0.00
21	Hexachlorobenzene	0.206	0.167	18.9	142	0.00
22	Pentachlorophenol	0.061	0.064	-4.9	226#	0.00
23	Phenanthrene	1.188	0.914	23.1	134	0.00
24	Anthracene	1.124	0.839	25.4#	132	0.00
25 SURR	Fluoranthene-d10	5.336	3.861	27.6#	123	0.00
26	Fluoranthene	1.671	1.158	30.7#	122	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	128	0.00
28	Pyrene	1.357	1.089	19.7	121	0.00
29 S	Terphenyl-d14	0.846	0.706	16.5	121	0.00
30	Benzo(a)anthracene	1.409	1.077	23.6	115	0.00
31	Chrysene	1.408	1.117	20.7	121	0.00
32	Bis(2-ethylhexyl)phthalate	1.829	1.619	11.5	160#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.514	1.199	20.8	120	0.00
34 I	Perylene-d12	1.000	1.000	0.0	134	0.00
35	Benzo(b)fluoranthene	1.384	1.024	26.0#	120	0.00
36	Benzo(k)fluoranthene	1.356	1.017	25.0#	120	0.00
37 C	Benzo(a)pyrene	1.279	0.953	25.5#	121	0.00
38	Dibenzo(a,h)anthracene	1.244	0.936	24.8	123	-0.01
39	Benzo(g,h,i)perylene	1.278	0.950	25.7#	119	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032219\
Data File : BE099101.D
Acq On : 23 Mar 2019 4:18
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 23 07:08:46 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 23 04:40:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 1		