

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032219\
 Data File : BE099081.D
 Acq On : 22 Mar 2019 14:34
 Operator : JU/SJ
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Manual Integrations
 APPROVED

Sohil
 3/25/2019 12:11:48 PM

Quant Time: Mar 22 15:35:15 2019

Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Fri Mar 22 13:31:55 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	3257	0.40	ng	0.00
7) Naphthalene-d8	10.64	136	13945	0.40	ng	0.00
13) Acenaphthene-d10	14.49	164	9915	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	28477	0.40	ng	0.00
27) Chrysene-d12	21.39	240	33458	0.40	ng	0.00
34) Perylene-d12	23.90	264	33007	0.40	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.41	112	47837	4.91	ng	0.00
5) Phenol-d6	7.00	99	70186	5.10	ng	0.00
8) Nitrobenzene-d5	8.99	82	36301	2.48	ng	0.00
11) 2-Methylnaphthalene-d10	12.22	152	123119	4.78	ng	-0.01
14) 2,4,6-Tribromophenol	0.00	330	0d	0.00	ng	
15) 2-Fluorobiphenyl	13.12	172	185334	4.51	ng	0.00
25) Fluoranthene-d10	19.25	212	1812340	3.99	ng	0.00
29) Terphenyl-d14	19.84	244	356657	4.39	ng	0.00

Target Compounds

					Qvalue
2) 1,4-Dioxane	3.35	88	23720	3.693 ng	92
3) n-Nitrosodimethylamine	3.66	42	16154m	1.971 ng	
6) bis(2-Chloroethyl)ether	7.26	93	59948	5.647 ng	100
9) Naphthalene	10.68	128	775951	4.858 ng	100
10) Hexachlorobutadiene	10.97	225	37344	3.534 ng	99
12) 2-Methylnaphthalene	12.31	142	137109	5.293 ng	99
16) Acenaphthylene	14.21	152	266521	5.227 ng	100
17) Acenaphthene	14.54	154	152223	5.032 ng	96
18) Fluorene	15.51	166	218456	5.008 ng	99
20) 4-Bromophenyl-phenylether	16.41	248	78087	5.059 ng	96
21) Hexachlorobenzene	16.52	284	71405	3.891 ng	98
22) Pentachlorophenol	16.87	266	29427	7.892 ng	98
23) Phenanthrene	17.25	178	397113	4.906 ng	99
24) Anthracene	17.35	178	393876	5.693 ng	100
26) Fluoranthene	19.28	202	540717	4.733 ng	99
28) Pyrene	19.64	202	551921	5.392 ng	100
30) Benzo(a)anthracene	21.38	228	557092	5.335 ng	98
31) Chrysene	21.43	228	545303	4.881 ng	99
32) Bis(2-ethylhexyl)phthalate	21.31	149	828133	3.727 ng	100
33) Indeno(1,2,3-cd)pyrene	26.57	276	613065	5.198 ng	99
35) Benzo(b)fluoranthene	23.13	252	544695	5.019 ng	98
36) Benzo(k)fluoranthene	23.18	252	533457	4.695 ng	# 97
37) Benzo(a)pyrene	23.79	252	504702	4.850 ng	95
38) Dibenzo(a,h)anthracene	26.59	278	505232	5.132 ng	99
39) Benzo(g,h,i)perylene	27.39	276	502490	4.880 ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 0.400 ng

RT: 7.84 min Scan# 432

Delta R.T. -0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Instrument :

BNA_E





ClientSampled :

SSTDICC5.0

Tgt Ion:	152	Resp:	3257
Ion Ratio	Lower	Upper	
152	100		
150	151.7	119.7	179.5
115	69.7	56.5	84.7

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:48 PM



#2

1,4-Dioxane

Concen: 3.693 ng




RT: 3.35 min Scan# 42




Delta R.T. -0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Tgt Ion:	88	Resp:	23720
Ion Ratio	Lower	Upper	
88	100		
58	63.6	45.5	68.3
43	20.5	14.9	22.3









#3
n-Nitrosodimethylamine
Concen: 1.971 ng m
RT: 3.66 min Scan# 69
Delta R.T. -0.03 min
Lab File: BE099081.D
Acq: 22 Mar 2019 14:34

Tgt Ion	Ratio	Resp	Lower	Upper
42	100	16154		
74	176.3	0.0	0.0	0.0
44	9.3	0.0	0.0	0.0

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0


Manual Integrations
APPROVED




Sohil
3/25/2019 12:11:48 PM



#4
2-Fluorophenol
Concen: 4.913 ng
RT: 5.41 min Scan# 221
Delta R.T. -0.00 min
Lab File: BE099081.D
Acq: 22 Mar 2019 14:34

Tgt Ion	Ratio	Resp	Lower	Upper
112	100	47837		
64	58.0	44.8	67.2	
63	118.1	88.2	132.2	





#5

Phenol-d6

Concen: 5.103 ng

RT: 7.00 min Scan# 359

Delta R.T. 0.01 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Tgt Ion:	99	Resp:	70186
Ion	Ratio	Lower	Upper
99	100		
42	14.3	11.8	17.8
71	35.9	27.4	41.2

Instrument :





BNA_E

ClientSampleId :

SSTDICC5.0

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:48 PM



#6

bis(2-Chloroethyl)ether

Concen: 5.647 ng


RT: 7.26 min Scan# 382


Delta R.T. -0.01 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Tgt Ion:	93	Resp:	59948
Ion	Ratio	Lower	Upper
93	100		
63	288.0	230.3	345.5
95	32.6	27.4	41.2





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.64 min Scan# 652

Delta R.T. 0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Instrument :

BNA_E


ClientSampleId :

SSTDICC5.0

Tgt Ion:	136	Resp:	13945
Ion Ratio	Lower	Upper	
136	100		
137	11.3	8.8	13.2
54	19.1	15.7	23.5
68	5.4	4.3	6.5

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:48 PM






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 2.483 ng


RT: 8.99 min Scan# 525

Delta R.T. 0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34



Tgt Ion:	82	Resp:	36301
Ion Ratio	Lower	Upper	
82	100		
128	104.7	98.8	148.2
54	158.8	143.7	215.5




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 4.858 ng

RT: 10.68 min Scan# 655

Delta R.T. 0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tgt Ion:	128	Resp:	775951
Ion Ratio	Lower	Upper	
128	100		
129	2.7	2.3	3.5
127	3.4	2.8	4.2

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:48 PM



#10

Hexachlorobutadiene

Concen: 3.534 ng




RT: 10.97 min Scan# 677


Delta R.T. 0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Tgt Ion:	225	Resp:	37344
Ion Ratio	Lower	Upper	
225	100		
223	63.4	51.1	76.7
227	63.9	51.4	77.2









#11
 2-Methylnaphthalene-d10
 Concen: 4.777 ng
 RT: 12.22 min Scan# 764
 Delta R.T. -0.01 min
 Lab File: BE099081.D
 Acq: 22 Mar 2019 14:34

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.7	23.5




Manual Integrations
 APPROVED

Sohil
 3/25/2019 12:11:48 PM



#12
 2-Methylnaphthalene
 Concen: 5.293 ng
 RT: 12.31 min Scan# 770
 Delta R.T. -0.00 min
 Lab File: BE099081.D
 Acq: 22 Mar 2019 14:34

Tgt Ion	Ratio	Lower	Upper
142	100		
141	85.6	69.7	104.5
115	35.0	28.2	42.4









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.49 min Scan# 921
 Delta R.T. -0.00 min
 Lab File: BE099081.D
 Acq: 22 Mar 2019 14:34

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tgt Ion	Ratio	Lower	Upper
164	100		
162	98.2	76.8	115.2
160	46.6	36.1	54.1

Manual Integrations
 APPROVED


Sohil
 3/25/2019 12:11:48 PM




#14
 2,4,6-Tribromophenol
 Concen: 0.000 ng
 Expected RT: 15.96 min

Lab File: BE099081.D
 Acq: 22 Mar 2019 14:34

Tgt Ion	Sig	Exp Ratio
330	100	
332	103.7	
141	38.4	









#15
2-Fluorobiphenyl
Concen: 4.515 ng
RT: 13.12 min Scan# 826
Delta R.T. -0.00 min
Lab File: BE099081.D
Acq: 22 Mar 2019 14:34

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.2	26.7	40.1
170	23.0	18.1	27.1




Manual Integrations
APPROVED


Sohil
3/25/2019 12:11:48 PM



#16
Acenaphthylene
Concen: 5.227 ng
RT: 14.21 min Scan# 902
Delta R.T. -0.00 min
Lab File: BE099081.D
Acq: 22 Mar 2019 14:34

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.7	23.5
153	12.9	10.2	15.4





#17

Acenaphthene

Concen: 5.032 ng

RT: 14.54 min Scan# 925

Delta R.T. -0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Instrument :

BNA_E


ClientSampleId :

SSTDICC5.0




Tgt Ion:	154	Resp:	152223
Ion Ratio	Lower	Upper	
154	100		
153	113.3	94.9	142.3
152	55.1	45.8	68.8

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:48 PM



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E



#18

Fluorene

Concen: 5.008 ng

RT: 15.51 min Scan# 992




Delta R.T. 0.00 min


Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Tgt Ion:	166	Resp:	218456
Ion Ratio	Lower	Upper	
166	100		
165	100.3	79.4	119.2
167	13.3	10.8	16.2

Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.21 min Scan# 1119

Delta R.T. 0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Instrument :

BNA_E





ClientSampled :

SSTDICC5.0

Tgt Ion:	188	Resp:	28477
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	10.5	8.1	12.1

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:48 PM



#20

4-Bromophenyl-phenylether

Concen: 5.059 ng




RT: 16.41 min Scan# 1059


Delta R.T. 0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Tgt Ion:	248	Resp:	78087
Ion Ratio	Lower	Upper	
248	100		
250	94.4	72.0	108.0
141	57.9	47.8	71.8





#21

Hexachlorobenzene

Concen: 3.891 ng

RT: 16.52 min Scan# 1067

Delta R.T. 0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tgt Ion:	284	Resp:	71405
Ion Ratio		Lower	Upper
284	100		
142	33.3	28.5	42.7
249	34.9	28.2	42.4

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:48 PM



#22

Pentachlorophenol

Concen: 7.892 ng




RT: 16.87 min Scan# 1093


Delta R.T. 0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Tgt Ion:	266	Resp:	29427
Ion Ratio		Lower	Upper
266	100		
264	60.5	47.4	71.0
268	65.5	50.7	76.1









#23
Phenanthrene
Concen: 4.906 ng
RT: 17.25 min Scan# 1122
Delta R.T. 0.00 min
Lab File: BE099081.D
Acq: 22 Mar 2019 14:34

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.5	16.0	24.0
179	15.5	12.1	18.1




Manual Integrations
APPROVED


Sohil
3/25/2019 12:11:48 PM



#24
Anthracene
Concen: 5.693 ng
RT: 17.35 min Scan# 1129
Delta R.T. 0.00 min
Lab File: BE099081.D
Acq: 22 Mar 2019 14:34

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.1	15.0	22.6
179	15.5	12.4	18.6





#25

Fluoranthene-d10

Concen: 3.987 ng

RT: 19.25 min Scan# 1389

Delta R.T. 0.01 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tgt Ion:	212	Resp:	1812340
Ion Ratio	Lower	Upper	
212	100		
106	2.7	2.2	3.4
104	1.5	1.3	1.9

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:48 PM



#26

Fluoranthene

Concen: 4.733 ng




RT: 19.28 min Scan# 1394

Delta R.T. -0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Tgt Ion:	202	Resp:	540717
Ion Ratio	Lower	Upper	
202	100		
101	9.6	7.5	11.3
203	17.5	13.7	20.5





#27

Chrysene-d12

Concen: 0.400 ng

RT: 21.39 min Scan# 1632

Delta R.T. -0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tgt Ion	Ratio	Lower	Upper
240	100		
120	7.0	5.0	7.6
236	27.1	22.2	33.4

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:48 PM



#28

Pyrene

Concen: 5.392 ng




RT: 19.64 min Scan# 1457


Delta R.T. 0.01 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.4	16.1	24.1
203	17.6	14.0	21.0









#29
 Terphenyl-d14
 Concen: 4.386 ng
 RT: 19.84 min Scan# 1493
 Delta R.T. 0.01 min
 Lab File: BE099081.D
 Acq: 22 Mar 2019 14:34

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC5.0

Tgt Ion	Ratio	Resp Lower	Resp Upper
244	100		
212	29.8	26.9	40.3
122	7.2	7.8	11.8




Manual Integrations
 APPROVED


Sohil
 3/25/2019 12:11:48 PM



#30
 Benzo(a)anthracene
 Concen: 5.335 ng
 RT: 21.38 min Scan# 1631
 Delta R.T. 0.01 min
 Lab File: BE099081.D
 Acq: 22 Mar 2019 14:34

Tgt Ion	Ratio	Resp Lower	Resp Upper
228	100		
226	26.6	22.4	33.6
229	20.3	15.6	23.4





#31

Chrysene

Concen: 4.881 ng

RT: 21.43 min Scan# 1635

Delta R.T. -0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Instrument :

BNA_E





ClientSampled :

SSTDICC5.0

Tgt Ion:	228	Resp:	545303
Ion Ratio	Lower	Upper	
228	100		
226	30.8	24.3	36.5
229	19.8	15.5	23.3

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:48 PM



#32

Bis(2-ethylhexyl)phthalate

Concen: 3.727 ng




RT: 21.31 min Scan# 1625


Delta R.T. -0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Tgt Ion:	149	Resp:	828133
Ion Ratio	Lower	Upper	
149	100		
167	7.2	5.7	8.5
279	1.0	0.8	1.2





#33

Indeno(1,2,3-cd)pyrene

Concen: 5.198 ng

RT: 26.57 min Scan# 2953

Delta R.T. 0.01 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Instrument :

BNA_E





ClientSampleId :

SSTDICC5.0

Tgt Ion:	276	Resp:	613065
Ion Ratio	Lower	Upper	
276	100		
138	19.1	15.8	23.8
277	24.8	19.4	29.2

Manual Integrations
APPROVED

Sohil
3/25/2019 12:11:48 PM



#34

Perylene-d12

Concen: 0.400 ng




RT: 23.90 min Scan# 2209


Delta R.T. -0.00 min

Lab File: BE099081.D

Acq: 22 Mar 2019 14:34

Tgt Ion:	264	Resp:	33007
Ion Ratio	Lower	Upper	
264	100		
260	24.4	19.0	28.6
265	30.2	24.1	36.1









#35
Benzo(b)fluoranthene
Concen: 5.019 ng
RT: 23.13 min Scan# 1993
Delta R.T. 0.00 min
Lab File: BE099081.D
Acq: 22 Mar 2019 14:34

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.1	18.4	27.6
125	7.5	7.0	10.4




Manual Integrations
APPROVED


Sohil
3/25/2019 12:11:48 PM



#36
Benzo(k)fluoranthene
Concen: 4.695 ng
RT: 23.18 min Scan# 2007
Delta R.T. 0.00 min
Lab File: BE099081.D
Acq: 22 Mar 2019 14:34

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.2	18.6	28.0
125	7.7	7.8	11.6









#37
Benzo(a)pyrene
Concen: 4.850 ng
RT: 23.79 min Scan# 2178
Delta R.T. 0.00 min
Lab File: BE099081.D
Acq: 22 Mar 2019 14:34

Instrument :
BNA_E
ClientSampleId :
SSTDICC5.0

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.0	19.3	28.9
125	8.3	8.3	12.5




Manual Integrations
APPROVED


Sohil
3/25/2019 12:11:48 PM



#38
Dibenzo(a,h)anthracene
Concen: 5.132 ng
RT: 26.59 min Scan# 2960
Delta R.T. -0.00 min
Lab File: BE099081.D
Acq: 22 Mar 2019 14:34

Tgt Ion	Ratio	Lower	Upper
278	100		
139	12.3	10.6	16.0
279	23.8	19.4	29.2





#39

Benzo(g,h,i)perylene

Concen: 4.880 ng

RT: 27.39 min Scan# 3182

Delta R.T. -0.00 min

Lab File: BE099081.D


Acq: 22 Mar 2019 14:34

Instrument :

BNA_E

ClientSampleId :

SSTDICC5.0



Tgt Ion: 276 Resp: 502490

Ion Ratio Lower Upper

276 100

277 23.7 19.4 29.2

138 16.6 14.6 21.8

Manual Integrations

APPROVED

Sohil

3/25/2019 12:11:48 PM

