

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032319\
 Data File : BE099124.D
 Acq On : 23 Mar 2019 23:25
 Operator : JU/SJ
 Sample : K2014-11
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PST-026-B124-031919

Quant Time: Mar 24 03:41:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 23 04:40:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	3139	0.40	ng	0.00
7) Naphthalene-d8	10.63	136	11866	0.40	ng	-0.01
13) Acenaphthene-d10	14.49	164	8397	0.40	ng	0.00
19) Phenanthrene-d10	17.22	188	25462	0.40	ng	0.00
27) Chrysene-d12	21.39	240	39828	0.40	ng	0.00
34) Perylene-d12	23.90	264	43018	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	3180	0.34	ng	0.00
5) Phenol-d6	6.99	99	4487	0.33	ng	0.00
8) Nitrobenzene-d5	8.99	82	3099	0.60	ng	0.00
11) 2-Methylnaphthalene-d10	12.22	152	6435	0.30	ng	-0.01
14) 2,4,6-Tribromophenol	15.96	330	1298	0.53	ng	0.00
15) 2-Fluorobiphenyl	13.10	172	9713	0.29	ng	-0.01
25) Fluoranthene-d10	19.24	212	107195	0.32	ng	0.00
29) Terphenyl-d14	19.83	244	21135	0.25	ng	0.00


Target Compounds				Qvalue		
9) Naphthalene	10.68	128	3424	0.024	ng	# 94
12) 2-Methylnaphthalene	12.30	142	575	0.024	ng	# 83
16) Acenaphthylene	14.20	152	4072	0.091	ng	# 94
18) Fluorene	15.52	166	1786	0.046	ng	# 88
22) Pentachlorophenol	16.87	266	54	0.261	ng	95
23) Phenanthrene	17.26	178	44039	0.583	ng	98
24) Anthracene	17.35	178	4555	0.064	ng	98
26) Fluoranthene	19.27	202	87996	0.827	ng	99
28) Pyrene	19.63	202	107420	0.795	ng	95
30) Benzo(a)anthracene	21.37	228	53561	0.382	ng	99
31) Chrysene	21.43	228	58169	0.415	ng	97
32) Bis(2-ethylhexyl)phthalate	21.29	149	44846	0.246	ng	98
33) Indeno(1,2,3-cd)pyrene	26.55	276	27626	0.183	ng	98
35) Benzo(b)fluoranthene	23.12	252	58141	0.391	ng	98
36) Benzo(k)fluoranthene	23.12	252	58141	0.399	ng	100
37) Benzo(a)pyrene	23.78	252	44972	0.327	ng	97
38) Dibenzo(a,h)anthracene	26.56	278	7589	0.057	ng	# 86
39) Benzo(g,h,i)perylene	27.37	276	31371	0.228	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032319\
Data File : BE099124.D
Acq On : 23 Mar 2019 23:25
Operator : JU/SJ
Sample : K2014-11
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PST-026-B124-031919

Quant Time: Mar 24 03:41:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 23 04:40:11 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.84 min Scan# 432
Delta R.T. -0.00 min
Lab File: BE099124.D
Acq: 23 Mar 2019 23:25

Instrument :
BNA_E
ClientSampleId :
PST-026-B124-031919

Tgt Ion:152 Resp: 3139
Ion Ratio Lower Upper
152 100
150 154.0 122.5 183.7
115 67.6 57.4 86.2




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#4


2-Fluorophenol
Concen: 0.340 ng
RT: 5.41 min Scan# 221
Delta R.T. -0.00 min
Lab File: BE099124.D
Acq: 23 Mar 2019 23:25

Tgt Ion:112 Resp: 3180
Ion Ratio Lower Upper
112 100
64 57.7 57.9 86.9#
63 121.0 158.3 237.5#



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BE0
Ion 63.00 (62.70 to 63.70): BE0





#5

Phenol-d6

Concen: 0.333 ng

RT: 6.99 min Scan# 358

Delta R.T. -0.00 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25

Instrument :

BNA_E

ClientSampleId :

PST-026-B124-031919


Tgt Ion: 99 Resp: 4487

Ion Ratio Lower Upper

99 100

42 0.0 24.2 36.2#




71 37.8 35.7 53.5



Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0



#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.63 min Scan# 651

Delta R.T. -0.01 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25

Tgt Ion: 136 Resp: 11866


Ion Ratio Lower Upper

136 100

137 11.8 11.4 17.0

54 32.9 39.3 58.9#

68 7.6 6.7 10.1






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0





#8

Nitrobenzene-d5

Concen: 0.597 ng

RT: 8.99 min Scan# 525

Delta R.T. -0.01 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25

Instrument :

BNA_E

ClientSampleId :

PST-026-B124-031919


Tgt Ion: 82 Resp: 3099

Ion Ratio Lower Upper

82 100

128 125.4 123.8 185.8




54 148.6 203.4 305.2#



Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0



#9

Naphthalene

Concen: 0.024 ng

RT: 10.68 min Scan# 655

Delta R.T. -0.00 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25


Tgt Ion: 128 Resp: 3424

Ion Ratio Lower Upper

128 100

129 5.7 2.6 4.0#



127 5.5 3.1 4.7#




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E









#11
 2-Methylnaphthalene-d10
 Concen: 0.301 ng
 RT: 12.22 min Scan# 764
 Delta R.T. -0.01 min
 Lab File: BE099124.D
 Acq: 23 Mar 2019 23:25




Instrument :
 BNA_E
 ClientSampleId :
 PST-026-B124-031919


Tgt Ion:152 Resp: 6435
 Ion Ratio Lower Upper
 152 100
 151 19.6 16.5 24.7



#12
 2-Methylnaphthalene
 Concen: 0.024 ng
 RT: 12.30 min Scan# 769
 Delta R.T. -0.01 min
 Lab File: BE099124.D
 Acq: 23 Mar 2019 23:25

Tgt Ion:142 Resp: 575
 Ion Ratio Lower Upper
 142 100
 141 106.5 72.6 108.8
 115 58.3 37.4 56.2#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.49 min Scan# 921
Delta R.T. -0.00 min
Lab File: BE099124.D
Acq: 23 Mar 2019 23:25




Instrument :
BNA_E
ClientSampleId :
PST-026-B124-031919


Tgt Ion:164 Resp: 8397
Ion Ratio Lower Upper
164 100
162 95.3 80.2 120.2
160 44.1 37.8 56.8



#14
2,4,6-Tribromophenol
Concen: 0.526 ng
RT: 15.96 min Scan# 1025
Delta R.T. 0.00 min
Lab File: BE099124.D
Acq: 23 Mar 2019 23:25

Tgt Ion:330 Resp: 1298
Ion Ratio Lower Upper
330 100
332 94.1 77.6 116.4
141 37.1 31.0 46.4









#15
2-Fluorobiphenyl
Concen: 0.291 ng
RT: 13.10 min Scan# 825
Delta R.T. -0.01 min
Lab File: BE099124.D
Acq: 23 Mar 2019 23:25

Instrument :
BNA_E
ClientSampleId :
PST-026-B124-031919

Tgt Ion:172 Resp: 9713
Ion Ratio Lower Upper
172 100
171 35.7 28.0 42.0
170 25.5 17.7 26.5




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#16
Acenaphthylene
Concen: 0.091 ng
RT: 14.20 min Scan# 901
Delta R.T. -0.01 min
Lab File: BE099124.D
Acq: 23 Mar 2019 23:25

Tgt Ion:152 Resp: 4072
Ion Ratio Lower Upper
152 100
151 16.9 17.0 25.6#
153 13.9 11.2 16.8



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E









#18
 Fluorene
 Concen: 0.046 ng
 RT: 15.52 min Scan# 992
 Delta R.T. 0.00 min
 Lab File: BE099124.D
 Acq: 23 Mar 2019 23:25

Instrument :
 BNA_E
 ClientSampleId :
 PST-026-B124-031919

Tgt Ion	Ratio	Lower	Upper
166	100		
165	109.7	79.3	118.9
167	22.8	10.5	15.7#




Abundance Ion 166.00 (165.70 to 166.70): E
 Ion 165.00 (164.70 to 165.70): E
 Ion 167.00 (166.70 to 167.70): E






#19
 Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.22 min Scan# 1119
 Delta R.T. 0.00 min
 Lab File: BE099124.D
 Acq: 23 Mar 2019 23:25

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.0	6.3	9.5#



Abundance Ion 188.00 (187.70 to 188.70): E
 Ion 94.00 (93.70 to 94.70): BE0
 Ion 80.00 (79.70 to 80.70): BE0





#22

Pentachlorophenol

Concen: 0.261 ng

RT: 16.87 min Scan# 1093

Delta R.T. 0.00 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25

Instrument :

BNA_E

ClientSampleId :

PST-026-B124-031919


Tgt Ion:266 Resp: 54

Ion Ratio Lower Upper

266 100

264 84.5 66.6 100.0

268 95.8 69.8 104.8






Abundance

Ion 266.00 (265.70 to 266.70): E

Ion 264.00 (263.70 to 264.70): E

Ion 268.00 (267.70 to 268.70): E



#23

Phenanthrene

Concen: 0.583 ng

RT: 17.26 min Scan# 1122

Delta R.T. 0.00 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25


Tgt Ion:178 Resp: 44039

Ion Ratio Lower Upper

178 100

176 19.2 16.0 24.0

179 15.7 12.1 18.1






Abundance

Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#24

Anthracene

Concen: 0.064 ng

RT: 17.35 min Scan# 1129

Delta R.T. 0.00 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25

Instrument :

BNA_E

ClientSampleId :

PST-026-B124-031919




Tgt Ion:178 Resp: 4555

Ion Ratio Lower Upper

178 100

176 18.7 15.4 23.0

179 17.0 12.2 18.4



#25

Fluoranthene-d10

Concen: 0.316 ng

RT: 19.24 min Scan# 1388

Delta R.T. -0.00 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25




Tgt Ion:212 Resp: 107195


Ion Ratio Lower Upper

212 100

106 2.8 1.8 2.8#

104 1.6 1.0 1.6#









#26
Fluoranthene
Concen: 0.827 ng
RT: 19.27 min Scan# 1393
Delta R.T. -0.01 min
Lab File: BE099124.D
Acq: 23 Mar 2019 23:25




Instrument :
BNA_E
ClientSampleId :
PST-026-B124-031919


Tgt Ion: 202 Resp: 87996
Ion Ratio Lower Upper
202 100
101 9.1 7.3 10.9
203 17.6 13.4 20.2



#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.39 min Scan# 1632
Delta R.T. -0.00 min
Lab File: BE099124.D
Acq: 23 Mar 2019 23:25

Tgt Ion: 240 Resp: 39828
Ion Ratio Lower Upper
240 100
120 3.7 5.0 7.4#
236 25.9 22.7 34.1





#28

Pyrene

Concen: 0.795 ng

RT: 19.63 min Scan# 1456

Delta R.T. -0.00 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25

Instrument :

BNA_E

ClientSampleId :

PST-026-B124-031919


Tgt Ion:202 Resp: 107420

Ion Ratio Lower Upper

202 100

200 19.8 17.0 25.6




203 20.4 14.2 21.2



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#29

Terphenyl-d14

Concen: 0.251 ng

RT: 19.83 min Scan# 1491

Delta R.T. -0.01 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25


Tgt Ion:244 Resp: 21135

Ion Ratio Lower Upper

244 100

212 31.5 29.4 44.2



122 9.4 7.1 10.7




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#30

Benzo(a)anthracene

Concen: 0.382 ng

RT: 21.37 min Scan# 1630

Delta R.T. -0.00 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25

Instrument :

BNA_E

ClientSampleId :

PST-026-B124-031919


Tgt Ion:228 Resp: 53561

Ion Ratio Lower Upper

228 100

226 28.7 23.0 34.4

229 21.9 16.5 24.7






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#31

Chrysene

Concen: 0.415 ng

RT: 21.43 min Scan# 1635

Delta R.T. -0.00 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25


Tgt Ion:228 Resp: 58169

Ion Ratio Lower Upper

228 100

226 29.2 23.7 35.5

229 22.6 16.1 24.1






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#32
 Bis(2-ethylhexyl)phthalate
 Concen: 0.246 ng
 RT: 21.29 min Scan# 1624
 Delta R.T. -0.01 min
 Lab File: BE099124.D
 Acq: 23 Mar 2019 23:25




Instrument :
 BNA_E
 ClientSampleId :
 PST-026-B124-031919


Tgt Ion: 149 Resp: 44846
 Ion Ratio Lower Upper
 149 100
 167 7.5 5.4 8.0
 279 1.0 1.0 1.4



#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.183 ng
 RT: 26.55 min Scan# 2947
 Delta R.T. -0.01 min
 Lab File: BE099124.D
 Acq: 23 Mar 2019 23:25

Tgt Ion: 276 Resp: 27626
 Ion Ratio Lower Upper
 276 100
 138 15.6 13.8 20.8
 277 24.7 19.5 29.3









#34
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.90 min Scan# 2208
 Delta R.T. -0.00 min
 Lab File: BE099124.D
 Acq: 23 Mar 2019 23:25




Instrument :
 BNA_E
 ClientSampleId :
 PST-026-B124-031919


Tgt Ion: 264 Resp: 43018
 Ion Ratio Lower Upper
 264 100
 260 23.3 21.2 31.8
 265 27.4 24.6 36.8



#35
 Benzo(b)fluoranthene
 Concen: 0.391 ng
 RT: 23.12 min Scan# 1991
 Delta R.T. -0.00 min
 Lab File: BE099124.D
 Acq: 23 Mar 2019 23:25

Tgt Ion: 252 Resp: 58141
 Ion Ratio Lower Upper
 252 100
 253 24.1 18.2 27.2
 125 9.5 7.3 10.9





#36

Benzo(k)fluoranthene

Concen: 0.399 ng

RT: 23.12 min Scan# 1991

Delta R.T. -0.05 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25

Instrument :

BNA_E

ClientSampleId :

PST-026-B124-031919


Tgt Ion:252 Resp: 58141

Ion Ratio Lower Upper

252 100

253 24.1 19.3 28.9

125 9.5 7.5 11.3






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#37

Benzo(a)pyrene

Concen: 0.327 ng

RT: 23.78 min Scan# 2175

Delta R.T. -0.01 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25


Tgt Ion:252 Resp: 44972

Ion Ratio Lower Upper

252 100

253 23.7 19.9 29.9

125 9.5 9.0 13.4






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#38

Dibenzo(a,h)anthracene

Concen: 0.057 ng

RT: 26.56 min Scan# 2951

Delta R.T. -0.03 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25

Instrument :

BNA_E

ClientSampleId :

PST-026-B124-031919


Tgt Ion: 278 Resp: 7589

Ion Ratio Lower Upper

278 100

139 18.9 11.8 17.6#




279 34.9 21.1 31.7#



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#39

Benzo(g,h,i)perylene

Concen: 0.228 ng

RT: 27.37 min Scan# 3177

Delta R.T. -0.02 min

Lab File: BE099124.D

Acq: 23 Mar 2019 23:25


Tgt Ion: 276 Resp: 31371

Ion Ratio Lower Upper

276 100

277 24.7 20.4 30.6

138 16.0 13.6 20.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

