

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032319\
 Data File : BE099126.D
 Acq On : 24 Mar 2019 00:36
 Operator : JU/SJ
 Sample : K2014-17
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PST-026-SW044-031919

Quant Time: Mar 24 03:48:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 23 04:40:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	3236	0.40	ng	0.00
7) Naphthalene-d8	10.63	136	13170	0.40	ng	-0.01
13) Acenaphthene-d10	14.49	164	9286	0.40	ng	0.00
19) Phenanthrene-d10	17.22	188	28470	0.40	ng	0.00
27) Chrysene-d12	21.38	240	39319	0.40	ng	-0.01
34) Perylene-d12	23.89	264	42286	0.40	ng	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.41	112	3355	0.35	ng	0.00
5) Phenol-d6	6.99	99	4711	0.34	ng	0.00
8) Nitrobenzene-d5	8.99	82	3345	0.58	ng	0.00
11) 2-Methylnaphthalene-d10	12.22	152	7241	0.31	ng	-0.01
14) 2,4,6-Tribromophenol	15.96	330	1476	0.54	ng	0.00
15) 2-Fluorobiphenyl	13.10	172	11574	0.31	ng	-0.01
25) Fluoranthene-d10	19.24	212	125650	0.33	ng	0.00
29) Terphenyl-d14	19.83	244	25368	0.30	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
23) Phenanthrene	17.26	178	3187	0.038	ng	# 92
26) Fluoranthene	19.27	202	6036	0.051	ng	# 96
28) Pyrene	19.63	202	7115	0.053	ng	# 96
30) Benzo(a)anthracene	21.37	228	3425	0.025	ng	# 90
31) Chrysene	21.43	228	4332	0.031	ng	# 93
32) Bis(2-ethylhexyl)phthalate	21.29	149	17191	0.096	ng	# 99
35) Benzo(b)fluoranthene	23.12	252	4967	0.034	ng	# 61
37) Benzo(a)pyrene	23.78	252	3480	0.026	ng	# 67
39) Benzo(g,h,i)perylene	27.37	276	2802	0.021	ng	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032319\
Data File : BE099126.D
Acq On : 24 Mar 2019 00:36
Operator : JU/SJ
Sample : K2014-17
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PST-026-SW044-031919

Quant Time: Mar 24 03:48:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 23 04:40:11 2019
Response via : Initial Calibration

