

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032319\
 Data File : BE099133.D
 Acq On : 24 Mar 2019 4:59
 Operator : JU/SJ
 Sample : K2014-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PST-026-SW075-031819

Manual Integrations
 APPROVED

Sohil
 3/25/2019 12:37:38 PM

Quant Time: Mar 24 23:53:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 23 04:40:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.84	152	3706	0.40	ng	0.00
7) Naphthalene-d8	10.63	136	14714	0.40	ng	-0.01
13) Acenaphthene-d10	14.49	164	9720	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	28342	0.40	ng	0.00
27) Chrysene-d12	21.38	240	39176	0.40	ng	-0.01
34) Perylene-d12	23.89	264	42649	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.41	112	4825	0.44	ng	0.00
5) Phenol-d6	6.99	99	6932	0.44	ng	0.00
8) Nitrobenzene-d5	8.99	82	5241	0.81	ng	-0.01
11) 2-Methylnaphthalene-d10	12.22	152	8966	0.34	ng	-0.01
14) 2,4,6-Tribromophenol	15.96	330	2072	0.73	ng	0.00
15) 2-Fluorobiphenyl	13.10	172	13918	0.36	ng	-0.01
25) Fluoranthene-d10	19.24	212	130502	0.35	ng	0.00
29) Terphenyl-d14	19.83	244	26045	0.31	ng	0.00

Target Compounds				Qvalue		
16) Acenaphthylene	14.20	152	2246	0.043	ng	97
23) Phenanthrene	17.25	178	18817	0.224	ng	97
24) Anthracene	17.35	178	2182	0.027	ng	# 92
26) Fluoranthene	19.27	202	39678	0.335	ng	98
28) Pyrene	19.63	202	48744	0.367	ng	97
30) Benzo(a)anthracene	21.37	228	24163	0.175	ng	96
31) Chrysene	21.42	228	24732	0.179	ng	93
32) Bis(2-ethylhexyl)phthalate	21.29	149	31624	0.177	ng	# 97
33) Indeno(1,2,3-cd)pyrene	26.55	276	13422	0.091	ng	97
35) Benzo(b)fluoranthene	23.12	252	26870	0.182	ng	# 85
36) Benzo(k)fluoranthene	23.17	252	9910m	0.069	ng	
37) Benzo(a)pyrene	23.78	252	21187	0.155	ng	# 97
38) Dibenzo(a,h)anthracene	26.57	278	3801	0.029	ng	# 55
39) Benzo(g,h,i)perylene	27.36	276	15264	0.112	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032319\
Data File : BE099133.D
Acq On : 24 Mar 2019 4:59
Operator : JU/SJ
Sample : K2014-01
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PST-026-SW075-031819

Manual Integrations
APPROVED

Sohil
3/25/2019 12:37:38 PM

Quant Time: Mar 24 23:53:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 23 04:40:11 2019
Response via : Initial Calibration

