

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032319\
 Data File : BE099103.D
 Acq On : 23 Mar 2019 9:30
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Mar 23 22:34:47 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 23 04:40:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.621	0.766	-23.3	130	-0.01
3	n-Nitrosodimethylamine	0.320	0.295	7.8	91	-0.01
4 S	2-Fluorophenol	1.191	1.389	-16.6	120	0.00
5 S	Phenol-d6	1.715	2.089	-21.8	127	0.00
6	bis(2-Chloroethyl)ether	1.394	1.646	-18.1	121	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.175	0.316	-80.6#	230#	-0.01
9	Naphthalene	4.714	5.193	-10.2	126	0.00
10	Hexachlorobutadiene	0.231	0.262	-13.4	128	0.00
11 SURR	2-Methylnaphthalene-d10	0.720	0.824	-14.4	132	-0.02
12	2-Methylnaphthalene	0.807	0.935	-15.9	136	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.118	0.188	-59.3#	247#	0.00
15 S	2-Fluorobiphenyl	1.591	1.730	-8.7	142	-0.02
16	Acenaphthylene	2.135	2.162	-1.3	143	-0.02
17	Acenaphthene	1.266	1.350	-6.6	149	0.00
18	Fluorene	1.847	1.959	-6.1	148	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
20	4-Bromophenyl-phenylether	0.221	0.253	-14.5	155#	0.00
21	Hexachlorobenzene	0.206	0.242	-17.5	162#	0.00
22	Pentachlorophenol	0.061	0.074	-21.3	203#	0.00
23	Phenanthrene	1.188	1.274	-7.2	147	0.00
24	Anthracene	1.124	1.137	-1.2	141	0.00
25 SURR	Fluoranthene-d10	5.336	4.972	6.8	124	0.00
26	Fluoranthene	1.671	1.551	7.2	129	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	101	-0.01
28	Pvrene	1.357	1.444	-6.4	127	0.00
29 S	Terphenyl-d14	0.846	0.996	-17.7	135	0.00
30	Benzo(a)anthracene	1.409	1.424	-1.1	120	0.00
31	Chrysene	1.408	1.508	-7.1	128	0.00
32	Bis(2-ethylhexyl)phthalate	1.829	1.611	11.9	125	-0.01
33	Indeno(1,2,3-cd)pyrene	1.514	1.580	-4.4	125	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.384	1.437	-3.8	126	0.00
36	Benzo(k)fluoranthene	1.356	1.392	-2.7	123	0.00
37 C	Benzo(a)pyrene	1.279	1.296	-1.3	123	0.00
38	Dibenzo(a,h)anthracene	1.244	1.293	-3.9	127	-0.01
39	Benzo(a,h,i)perylene	1.278	1.335	-4.5	126	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032319\
Data File : BE099103.D
Acq On : 23 Mar 2019 9:30
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 23 22:34:47 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 23 04:40:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			