

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032319\
 Data File : BE099120.D
 Acq On : 23 Mar 2019 20:08
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 22:24:19 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 23 04:40:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.621	0.657	-5.8	121	-0.01
3	n-Nitrosodimethylamine	0.320	0.357	-11.6	120	-0.01
4 S	2-Fluorophenol	1.191	1.419	-19.1	133	0.00
5 S	Phenol-d6	1.715	2.124	-23.8	141	0.00
6	bis(2-Chloroethyl)ether	1.394	1.621	-16.3	129	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	109	-0.01
8 S	Nitrobenzene-d5	0.175	0.366	-109.1#	286#	-0.01
9	Naphthalene	4.714	5.207	-10.5	136	0.00
10	Hexachlorobutadiene	0.231	0.267	-15.6	140	0.00
11 SURR	2-Methylnaphthalene-d10	0.720	0.851	-18.2	147	-0.02
12	2-Methylnaphthalene	0.807	0.953	-18.1	149	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	127	0.00
14 S	2,4,6-Tribromophenol	0.118	0.236	-100.0#	329#	0.00
15 S	2-Fluorobiphenyl	1.591	1.809	-13.7	158#	-0.02
16	Acenaphthylene	2.135	2.305	-8.0	162#	-0.02
17	Acenaphthene	1.266	1.381	-9.1	162#	0.00
18	Fluorene	1.847	2.033	-10.1	162#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	124	0.00
20	4-Bromophenyl-phenylether	0.221	0.269	-21.7	176#	0.00
21	Hexachlorobenzene	0.206	0.251	-21.8	180#	0.00
22	Pentachlorophenol	0.061	0.098	-60.7#	289#	-0.01
23	Phenanthrene	1.188	1.289	-8.5	159#	0.00
24	Anthracene	1.124	1.176	-4.6	156#	0.00
25 SURR	Fluoranthene-d10	5.336	5.395	-1.1	144	0.00
26	Fluoranthene	1.671	1.648	1.4	146	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	116	-0.01
28	Pvrene	1.357	1.442	-6.3	146	0.00
29 S	Terphenyl-d14	0.846	0.998	-18.0	156#	0.00
30	Benzo(a)anthracene	1.409	1.485	-5.4	144	0.00
31	Chrysene	1.408	1.500	-6.5	148	0.00
32	Bis(2-ethylhexyl)phthalate	1.829	2.234	-22.1	201#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.514	1.657	-9.4	151#	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	124	0.00
35	Benzo(b)fluoranthene	1.384	1.383	0.1	150	0.00
36	Benzo(k)fluoranthene	1.356	1.403	-3.5	154#	0.00
37 C	Benzo(a)pyrene	1.279	1.291	-0.9	152#	0.00
38	Dibenzo(a,h)anthracene	1.244	1.270	-2.1	154#	-0.02
39	Benzo(a,h,i)perylene	1.278	1.279	-0.1	149	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032319\
Data File : BE099120.D
Acq On : 23 Mar 2019 20:08
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 23 22:24:19 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 23 04:40:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			