

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032319\
 Data File : BE099122.D
 Acq On : 23 Mar 2019 22:01
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 23 22:43:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 23 04:40:11 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.621	0.654	-5.3	121	-0.01
3	n-Nitrosodimethylamine	0.320	0.339	-5.9	115	-0.01
4 S	2-Fluorophenol	1.191	1.381	-16.0	130	0.00
5 S	Phenol-d6	1.715	2.072	-20.8	138	0.00
6	bis(2-Chloroethyl)ether	1.394	1.672	-19.9	134	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	112	-0.01
8 S	Nitrobenzene-d5	0.175	0.339	-93.7#	272#	0.00
9	Naphthalene	4.714	5.202	-10.4	140	0.00
10	Hexachlorobutadiene	0.231	0.270	-16.9	146	0.00
11 SURR	2-Methylnaphthalene-d10	0.720	0.827	-14.9	147	-0.02
12	2-Methylnaphthalene	0.807	0.937	-16.1	150#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.00
14 S	2,4,6-Tribromophenol	0.118	0.227	-92.4#	326#	0.00
15 S	2-Fluorobiphenyl	1.591	1.768	-11.1	159#	-0.02
16	Acenaphthylene	2.135	2.256	-5.7	164#	-0.02
17	Acenaphthene	1.266	1.371	-8.3	166#	0.00
18	Fluorene	1.847	1.965	-6.4	162#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	126	0.00
20	4-Bromophenyl-phenylether	0.221	0.256	-15.8	172#	0.00
21	Hexachlorobenzene	0.206	0.251	-21.8	183#	0.00
22	Pentachlorophenol	0.061	0.085	-39.3#	257#	-0.01
23	Phenanthrene	1.188	1.260	-6.1	159#	0.00
24	Anthracene	1.124	1.156	-2.8	156#	0.00
25 SURR	Fluoranthene-d10	5.336	5.330	0.1	146	0.00
26	Fluoranthene	1.671	1.626	2.7	147	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	117	-0.01
28	Pvrene	1.357	1.435	-5.7	147	0.00
29 S	Terphenyl-d14	0.846	1.000	-18.2	158#	0.00
30	Benzo(a)anthracene	1.409	1.439	-2.1	141	0.00
31	Chrysene	1.408	1.506	-7.0	149	0.00
32	Bis(2-ethylhexyl)phthalate	1.829	1.964	-7.4	178#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.514	1.647	-8.8	152#	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	125	0.00
35	Benzo(b)fluoranthene	1.384	1.386	-0.1	151#	0.00
36	Benzo(k)fluoranthene	1.356	1.372	-1.2	151#	0.00
37 C	Benzo(a)pyrene	1.279	1.284	-0.4	152#	0.00
38	Dibenzo(a,h)anthracene	1.244	1.275	-2.5	155#	-0.02
39	Benzo(a,h,i)perylene	1.278	1.296	-1.4	152#	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032319\
Data File : BE099122.D
Acq On : 23 Mar 2019 22:01
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 23 22:43:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 23 04:40:11 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			