

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032519\
 Data File : BE099185.D
 Acq On : 25 Mar 2019 18:24
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 26 07:08:00 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	-0.01
2	1,4-Dioxane	0.527	0.475	9.9	96	0.00
3	n-Nitrosodimethylamine	0.295	0.200	32.2#	83	0.00
4 S	2-Fluorophenol	1.194	1.103	7.6	96	0.00
5 S	Phenol-d6	1.825	1.687	7.6	97	0.00
6	bis(2-Chloroethyl)ether	1.435	1.297	9.6	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
8 S	Nitrobenzene-d5	0.327	0.276	15.6	99	0.00
9	Naphthalene	4.717	4.190	11.2	96	0.00
10	Hexachlorobutadiene	0.251	0.223	11.2	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.771	0.677	12.2	94	0.00
12	2-Methylnaphthalene	0.843	0.738	12.5	94	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.141	0.087	38.3#	89	0.00
15 S	2-Fluorobiphenyl	1.631	1.467	10.1	94	0.00
16	Acenaphthylene	2.039	1.709	16.2	93	0.00
17	Acenaphthene	1.220	1.067	12.5	95	0.00
18	Fluorene	1.715	1.473	14.1	94	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.258	0.204	20.9	90	0.00
21	Hexachlorobenzene	0.249	0.202	18.9	91	0.00
22	Pentachlorophenol	0.033	0.017	48.5#	119	0.00
23	Phenanthrene	1.175	1.000	14.9	95	0.00
24	Anthracene	1.061	0.869	18.1	95	0.01
25 SURR	Fluoranthene-d10	4.847	4.196	13.4	96	0.00
26	Fluoranthene	1.499	1.249	16.7	95	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
28	Pvrene	1.329	1.134	14.7	96	0.00
29 S	Terphenyl-d14	0.870	0.727	16.4	91	0.00
30	Benzo(a)anthracene	1.320	1.085	17.8	96	0.00
31	Chrysene	1.411	1.207	14.5	97	0.01
32	Bis(2-ethylhexyl)phthalate	1.574	1.124	28.6#	99	0.00
33	Indeno(1,2,3-cd)pyrene	1.335	0.963	27.9#	89	0.00
34 I	Perylene-d12	1.000	1.000	0.0	90	0.00
35	Benzo(b)fluoranthene	1.316	1.048	20.4	87	0.00
36	Benzo(k)fluoranthene	1.392	1.209	13.1	92	0.00
37 C	Benzo(a)pyrene	1.229	1.034	15.9	92	0.00
38	Dibenzo(a,h)anthracene	1.092	0.797	27.0#	85	0.00
39	Benzo(a,h,i)perylene	1.095	0.814	25.7#	84	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032519\
Data File : BE099185.D
Acq On : 25 Mar 2019 18:24
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 26 07:08:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			