

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032519\
 Data File : BE099202.D
 Acq On : 26 Mar 2019 4:59
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 26 08:04:59 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.527	0.460	12.7	96	0.00
3	n-Nitrosodimethylamine	0.295	0.256	13.2	110	0.00
4 S	2-Fluorophenol	1.194	1.057	11.5	95	0.00
5 S	Phenol-d6	1.825	1.618	11.3	96	0.00
6	bis(2-Chloroethyl)ether	1.435	1.218	15.1	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.327	0.288	11.9	104	0.00
9	Naphthalene	4.717	4.197	11.0	97	0.00
10	Hexachlorobutadiene	0.251	0.221	12.0	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.771	0.701	9.1	98	0.00
12	2-Methylnaphthalene	0.843	0.752	10.8	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	-0.01
14 S	2,4,6-Tribromophenol	0.141	0.121	14.2	131	0.00
15 S	2-Fluorobiphenyl	1.631	1.475	9.6	100	0.00
16	Acenaphthylene	2.039	1.707	16.3	98	0.00
17	Acenaphthene	1.220	1.085	11.1	102	0.00
18	Fluorene	1.715	1.514	11.7	102	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4-Bromophenyl-phenylether	0.258	0.214	17.1	100	0.00
21	Hexachlorobenzene	0.249	0.207	16.9	99	-0.01
22	Pentachlorophenol	0.033	0.033	0.0	238#	0.00
23	Phenanthrene	1.175	1.013	13.8	101	0.00
24	Anthracene	1.061	0.877	17.3	102	0.00
25 SURR	Fluoranthene-d10	4.847	4.305	11.2	104	0.00
26	Fluoranthene	1.499	1.290	13.9	104	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	108	-0.01
28	Pvrene	1.329	1.116	16.0	104	0.00
29 S	Terphenyl-d14	0.870	0.715	17.8	98	0.00
30	Benzo(a)anthracene	1.320	1.109	16.0	108	0.00
31	Chrysene	1.411	1.192	15.5	105	0.00
32	Bis(2-ethylhexyl)phthalate	1.574	1.223	22.3	119	-0.01
33	Indeno(1,2,3-cd)pyrene	1.335	0.961	28.0#	97	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.316	1.074	18.4	99	0.00
36	Benzo(k)fluoranthene	1.392	1.255	9.8	106	0.00
37 C	Benzo(a)pyrene	1.229	1.041	15.3	103	0.00
38	Dibenzo(a,h)anthracene	1.092	0.808	26.0#	95	-0.02
39	Benzo(a,h,i)perylene	1.095	0.840	23.3	96	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032519\
Data File : BE099202.D
Acq On : 26 Mar 2019 4:59
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 26 08:04:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			