

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032519\
 Data File : BE099204.D
 Acq On : 26 Mar 2019 7:08
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 26 08:10:52 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	102	0.00
2	1,4-Dioxane	0.400	0.365	8.8	106	0.00
3	n-Nitrosodimethylamine	0.400	0.299	25.3#	100	0.00
4 S	2-Fluorophenol	0.400	0.343	14.2	96	0.00
5 S	Phenol-d6	0.400	0.354	11.5	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.351	12.3	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	99	0.00
8 S	Nitrobenzene-d5	0.400	0.333	16.8	101	0.00
9	Naphthalene	0.400	0.356	11.0	99	0.00
10	Hexachlorobutadiene	0.400	0.367	8.3	105	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.361	9.8	100	0.00
12	2-Methylnaphthalene	0.400	0.347	13.3	97	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.400	0.435	-8.7	127	0.00
15 S	2-Fluorobiphenyl	0.400	0.362	9.5	100	0.00
16	Acenaphthylene	0.400	0.331	17.3	98	0.00
17	Acenaphthene	0.400	0.350	12.5	100	0.00
18	Fluorene	0.400	0.348	13.0	101	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	103	0.00
20	4-Bromophenyl-phenylether	0.400	0.324	19.0	99	0.00
21	Hexachlorobenzene	0.400	0.329	17.8	100	0.00
22	Pentachlorophenol	0.400	0.596	-49.0#	247	0.00
23	Phenanthrene	0.400	0.338	15.5	101	0.00
24	Anthracene	0.400	0.327	18.3	103	0.01
25 SURR	Fluoranthene-d10	0.400	0.360	10.0	107	0.00
26	Fluoranthene	0.400	0.344	14.0	106	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	111	-0.01
28	Pvrene	0.400	0.337	15.8	108	0.00
29 S	Terphenyl-d14	0.400	0.330	17.5	101	0.00
30	Benzo(a)anthracene	0.400	0.333	16.8	110	0.00
31	Chrysene	0.400	0.336	16.0	108	0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.311	22.3	123	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.287	28.3#	100	0.00
34 I	Perylene-d12	0.400	0.400	0.0	102	0.00
35	Benzo(b)fluoranthene	0.400	0.327	18.3	101	0.00
36	Benzo(k)fluoranthene	0.400	0.365	8.8	109	0.00
37 C	Benzo(a)pyrene	0.400	0.346	13.5	107	0.00
38	Dibenzo(a,h)anthracene	0.400	0.290	27.5#	95	0.00
39	Benzo(a,h,i)perylene	0.400	0.305	23.8	98	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE032519\
Data File : BE099204.D
Acq On : 26 Mar 2019 7:08
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 26 08:10:52 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				