

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_E\DATA\BE032619\  
 Data File : BE099219.D  
 Acq On : 26 Mar 2019 17:34  
 Operator : JU/SJ  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Mar 27 09:09:14 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\8270-SIM-BE032519.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Mar 26 07:02:06 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	112	0.00
2	1,4-Dioxane	0.400	0.334	16.5	106	0.00
3	n-Nitrosodimethylamine	0.400	0.313	21.8	115	0.00
4 S	2-Fluorophenol	0.400	0.340	15.0	104	0.00
5 S	Phenol-d6	0.400	0.348	13.0	109	0.00
6	bis(2-Chloroethyl)ether	0.400	0.340	15.0	108	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	107	0.00
8 S	Nitrobenzene-d5	0.400	0.354	11.5	117	0.00
9	Naphthalene	0.400	0.357	10.8	109	0.00
10	Hexachlorobutadiene	0.400	0.363	9.3	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.364	9.0	110	0.00
12	2-Methylnaphthalene	0.400	0.355	11.3	108	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	111	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.447	-11.7	148	0.00
15 S	2-Fluorobiphenyl	0.400	0.359	10.3	110	0.00
16	Acenaphthylene	0.400	0.335	16.3	110	0.00
17	Acenaphthene	0.400	0.352	12.0	112	0.00
18	Fluorene	0.400	0.351	12.3	113	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	110	-0.01
20	4-Bromophenyl-phenylether	0.400	0.331	17.3	109	-0.01
21	Hexachlorobenzene	0.400	0.336	16.0	110	-0.01
22	Pentachlorophenol	0.400	0.678	-69.5#	331	-0.01
23	Phenanthrene	0.400	0.341	14.8	110	-0.01
24	Anthracene	0.400	0.336	16.0	113	0.00
25 SURR	Fluoranthene-d10	0.400	0.360	10.0	115	0.00
26	Fluoranthene	0.400	0.344	14.0	114	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	119	-0.01
28	Pyrene	0.400	0.337	15.8	116	0.00
29 S	Terphenyl-d14	0.400	0.333	16.8	110	0.00
30	Benzo(a)anthracene	0.400	0.337	15.8	120	0.00
31	Chrysene	0.400	0.337	15.8	116	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.321	19.8	136	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.301	24.8	112	0.00
34 I	Perylene-d12	0.400	0.400	0.0	112	0.00
35	Benzo(b)fluoranthene	0.400	0.328	18.0	112	0.00
36	Benzo(k)fluoranthene	0.400	0.358	10.5	118	0.00
37 C	Benzo(a)pyrene	0.400	0.340	15.0	115	0.00
38	Dibenzo(a,h)anthracene	0.400	0.315	21.3	114	-0.01
39	Benzo(g,h,i)perylene	0.400	0.301	24.8	106	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA\_E\DATA\BE032619\  
Data File : BE099219.D  
Acq On : 26 Mar 2019 17:34  
Operator : JU/SJ  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_E  
LabSampleId :  
SSTDCCC0.4

Quant Time: Mar 27 09:09:14 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_E\METHODS\8270-SIM-BE032519.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Mar 26 07:02:06 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
-----					
(#) = Out of Range	SPCC's out = 0		CCC's out = 0		