

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032619\
 Data File : BE099236.D
 Acq On : 27 Mar 2019 4:18
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 09:24:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	116	0.00
2	1,4-Dioxane	0.400	0.333	16.8	109	0.00
3	n-Nitrosodimethylamine	0.400	0.334	16.5	126	0.00
4 S	2-Fluorophenol	0.400	0.328	18.0	104	0.00
5 S	Phenol-d6	0.400	0.332	17.0	107	0.00
6	bis(2-Chloroethyl)ether	0.400	0.328	18.0	107	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	111	0.00
8 S	Nitrobenzene-d5	0.400	0.339	15.3	116	0.00
9	Naphthalene	0.400	0.346	13.5	109	-0.01
10	Hexachlorobutadiene	0.400	0.348	13.0	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.355	11.3	111	0.00
12	2-Methylnaphthalene	0.400	0.348	13.0	109	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	116	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.429	-7.2	142	0.00
15 S	2-Fluorobiphenyl	0.400	0.346	13.5	111	0.00
16	Acenaphthylene	0.400	0.320	20.0	109	0.00
17	Acenaphthene	0.400	0.334	16.5	111	0.00
18	Fluorene	0.400	0.336	16.0	113	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	115	-0.01
20	4-Bromophenyl-phenylether	0.400	0.322	19.5	110	-0.01
21	Hexachlorobenzene	0.400	0.325	18.8	111	-0.01
22	Pentachlorophenol	0.400	0.590	-47.5#	272	-0.01
23	Phenanthrene	0.400	0.329	17.8	111	-0.01
24	Anthracene	0.400	0.320	20.0	113	0.00
25 SURR	Fluoranthene-d10	0.400	0.345	13.8	116	0.00
26	Fluoranthene	0.400	0.329	17.8	114	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	122	-0.01
28	Pyrene	0.400	0.326	18.5	115	0.00
29 S	Terphenyl-d14	0.400	0.326	18.5	110	0.00
30	Benzo(a)anthracene	0.400	0.318	20.5	116	0.00
31	Chrysene	0.400	0.332	17.0	117	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.302	24.5	131	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.254	36.5#	97	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	108	0.00
35	Benzo(b)fluoranthene	0.400	0.321	19.8	106	0.00
36	Benzo(k)fluoranthene	0.400	0.352	12.0	112	0.00
37 C	Benzo(a)pyrene	0.400	0.336	16.0	110	0.00
38	Dibenzo(a,h)anthracene	0.400	0.240	40.0#	84	-0.02
39	Benzo(g,h,i)perylene	0.400	0.268	33.0#	91	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032619\
Data File : BE099236.D
Acq On : 27 Mar 2019 4:18
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 27 09:24:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				