

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032619\
 Data File : BE099236.D
 Acq On : 27 Mar 2019 4:18
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 09:24:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2	1,4-Dioxane	0.527	0.439	16.7	109	0.00
3	n-Nitrosodimethylamine	0.295	0.246	16.6	126	0.00
4 S	2-Fluorophenol	1.194	0.978	18.1	104	0.00
5 S	Phenol-d6	1.825	1.512	17.2	107	0.00
6	bis(2-Chloroethyl)ether	1.435	1.175	18.1	107	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	0.00
8 S	Nitrobenzene-d5	0.327	0.277	15.3	116	0.00
9	Naphthalene	4.717	4.076	13.6	109	-0.01
10	Hexachlorobutadiene	0.251	0.219	12.7	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.771	0.683	11.4	111	0.00
12	2-Methylnaphthalene	0.843	0.734	12.9	109	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	116	-0.01
14 S	2,4,6-Tribromophenol	0.141	0.113	19.9	142	0.00
15 S	2-Fluorobiphenyl	1.631	1.411	13.5	111	0.00
16	Acenaphthylene	2.039	1.630	20.1	109	0.00
17	Acenaphthene	1.220	1.018	16.6	111	0.00
18	Fluorene	1.715	1.442	15.9	113	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	115	-0.01
20	4-Bromophenyl-phenylether	0.258	0.208	19.4	110	-0.01
21	Hexachlorobenzene	0.249	0.203	18.5	111	-0.01
22	Pentachlorophenol	0.033	0.033	0.0	272#	-0.01
23	Phenanthrene	1.175	0.966	17.8	111	-0.01
24	Anthracene	1.061	0.849	20.0	113	0.00
25 SURR	Fluoranthene-d10	4.847	4.182	13.7	116	0.00
26	Fluoranthene	1.499	1.235	17.6	114	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	122	-0.01
28	Pyrene	1.329	1.083	18.5	115	0.00
29 S	Terphenyl-d14	0.870	0.709	18.5	110	0.00
30	Benzo(a)anthracene	1.320	1.049	20.5	116	0.00
31	Chrysene	1.411	1.170	17.1	117	0.00
32	Bis(2-ethylhexyl)phthalate	1.574	1.188	24.5	131	-0.01
33	Indeno(1,2,3-cd)pyrene	1.335	0.847	36.6#	97	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	108	0.00
35	Benzo(b)fluoranthene	1.316	1.057	19.7	106	0.00
36	Benzo(k)fluoranthene	1.392	1.224	12.1	112	0.00
37 C	Benzo(a)pyrene	1.229	1.031	16.1	110	0.00
38	Dibenzo(a,h)anthracene	1.092	0.656	39.9#	84	-0.02
39	Benzo(g,h,i)perylene	1.095	0.733	33.1#	91	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032619\
Data File : BE099236.D
Acq On : 27 Mar 2019 4:18
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 27 09:24:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		