

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032619\
 Data File : BE099238.D
 Acq On : 27 Mar 2019 6:51
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 14:41:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	-0.01
2	1,4-Dioxane	0.527	0.435	17.5	110	0.00
3	n-Nitrosodimethylamine	0.295	0.220	25.4#	115	0.00
4 S	2-Fluorophenol	1.194	0.984	17.6	106	-0.01
5 S	Phenol-d6	1.825	1.588	13.0	114	0.00
6	bis(2-Chloroethyl)ether	1.435	1.199	16.4	112	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
8 S	Nitrobenzene-d5	0.327	0.268	18.0	119	0.00
9	Naphthalene	4.717	4.051	14.1	114	0.00
10	Hexachlorobutadiene	0.251	0.216	13.9	117	0.00
11 SURR	2-Methylnaphthalene-d10	0.771	0.656	14.9	112	0.00
12	2-Methylnaphthalene	0.843	0.729	13.5	115	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	-0.01
14 S	2,4,6-Tribromophenol	0.141	0.111	21.3	147	0.00
15 S	2-Fluorobiphenyl	1.631	1.386	15.0	116	0.00
16	Acenaphthylene	2.039	1.626	20.3	115	0.00
17	Acenaphthene	1.220	1.025	16.0	118	0.00
18	Fluorene	1.715	1.425	16.9	118	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	121	-0.01
20	4-Bromophenyl-phenylether	0.258	0.206	20.2	115	-0.01
21	Hexachlorobenzene	0.249	0.209	16.1	120	-0.01
22	Pentachlorophenol	0.033	0.032	3.0	275#	-0.01
23	Phenanthrene	1.175	0.976	16.9	117	-0.01
24	Anthracene	1.061	0.832	21.6	116	0.00
25 SURR	Fluoranthene-d10	4.847	4.178	13.8	121	0.00
26	Fluoranthene	1.499	1.241	17.2	121	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	126	-0.01
28	Pyrene	1.329	1.120	15.7	122	0.00
29 S	Terphenyl-d14	0.870	0.725	16.7	116	0.00
30	Benzo(a)anthracene	1.320	1.070	18.9	122	0.00
31	Chrysene	1.411	1.191	15.6	123	0.00
32	Bis(2-ethylhexyl)phthalate	1.574	1.154	26.7#	131	-0.01
33	Indeno(1,2,3-cd)pyrene	1.335	0.766	42.6#	91	0.00
34 I	Perylene-d12	1.000	1.000	0.0	112	0.00
35	Benzo(b)fluoranthene	1.316	1.041	20.9	108	0.00
36	Benzo(k)fluoranthene	1.392	1.209	13.1	115	0.00
37 C	Benzo(a)pyrene	1.229	1.032	16.0	115	0.00
38	Dibenzo(a,h)anthracene	1.092	0.772	29.3#	103	-0.02
39	Benzo(g,h,i)perylene	1.095	0.657	40.0#	85	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032619\
Data File : BE099238.D
Acq On : 27 Mar 2019 6:51
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 27 14:41:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		