

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032619\
 Data File : BE099246.D
 Acq On : 27 Mar 2019 12:28
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 27 15:06:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	126	-0.01
2	1,4-Dioxane	0.400	0.323	19.3	115	0.00
3	n-Nitrosodimethylamine	0.400	0.321	19.8	132	-0.01
4 S	2-Fluorophenol	0.400	0.333	16.8	115	0.00
5 S	Phenol-d6	0.400	0.343	14.2	120	0.00
6	bis(2-Chloroethyl)ether	0.400	0.331	17.3	118	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	125	0.00
8 S	Nitrobenzene-d5	0.400	0.339	15.3	131	-0.01
9	Naphthalene	0.400	0.345	13.8	122	-0.01
10	Hexachlorobutadiene	0.400	0.340	15.0	123	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.354	11.5	124	-0.01
12	2-Methylnaphthalene	0.400	0.346	13.5	122	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	134	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.433	-8.2	168	0.00
15 S	2-Fluorobiphenyl	0.400	0.335	16.3	125	0.00
16	Acenaphthylene	0.400	0.323	19.3	128	0.00
17	Acenaphthene	0.400	0.338	15.5	130	-0.01
18	Fluorene	0.400	0.337	15.8	131	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	135	-0.01
20	4-Bromophenyl-phenylether	0.400	0.313	21.8	126	-0.01
21	Hexachlorobenzene	0.400	0.325	18.8	130	-0.01
22	Pentachlorophenol	0.400	0.522	-30.5#	254	-0.01
23	Phenanthrene	0.400	0.329	17.8	129	-0.01
24	Anthracene	0.400	0.315	21.3	130	0.00
25 SURR	Fluoranthene-d10	0.400	0.345	13.8	135	0.00
26	Fluoranthene	0.400	0.330	17.5	134	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	142	-0.01
28	Pyrene	0.400	0.331	17.3	136	0.00
29 S	Terphenyl-d14	0.400	0.328	18.0	129	0.00
30	Benzo(a)anthracene	0.400	0.316	21.0	134	0.00
31	Chrysene	0.400	0.325	18.8	134	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.333	16.8	168	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.221	44.8#	99	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	123	0.00
35	Benzo(b)fluoranthene	0.400	0.319	20.3	120	0.00
36	Benzo(k)fluoranthene	0.400	0.354	11.5	128	0.00
37 C	Benzo(a)pyrene	0.400	0.338	15.5	126	0.00
38	Dibenzo(a,h)anthracene	0.400	0.283	29.3#	112	-0.02
39	Benzo(g,h,i)perylene	0.400	0.250	37.5#	97	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE032619\
Data File : BE099246.D
Acq On : 27 Mar 2019 12:28
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 27 15:06:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				