

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE040219\
 Data File : BE099277.D
 Acq On : 2 Apr 2019 16:56
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE040219

Quant Time: Apr 02 19:23:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 02 17:12:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	-0.03
2	1,4-Dioxane	0.400	0.389	2.8	104	-0.03
3	n-Nitrosodimethylamine	0.400	0.403	-0.8	102	-0.03
4 S	2-Fluorophenol	0.400	0.408	-2.0	95	-0.03
5 S	Phenol-d6	0.400	0.390	2.5	94	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.403	-0.8	98	-0.03
7 I	Naphthalene-d8	0.400	0.400	0.0	96	-0.01
8 S	Nitrobenzene-d5	0.400	0.421	-5.2	105	-0.02
9	Naphthalene	0.400	0.401	-0.3	96	-0.02
10	Hexachlorobutadiene	0.400	0.404	-1.0	96	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.391	2.3	95	-0.01
12	2-Methylnaphthalene	0.400	0.384	4.0	93	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	96	-0.03
14 S	2,4,6-Tribromophenol	0.400	0.338	15.5	91	-0.01
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	95	-0.01
16	Acenaphthylene	0.400	0.366	8.5	91	-0.01
17	Acenaphthene	0.400	0.384	4.0	95	-0.01
18	Fluorene	0.400	0.376	6.0	90	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	94	-0.03
20	4-Bromophenyl-phenylether	0.400	0.387	3.3	93	-0.03
21	Hexachlorobenzene	0.400	0.409	-2.2	97	-0.03
22	Pentachlorophenol	0.400	0.325	18.8	87	-0.03
23	Phenanthrene	0.400	0.410	-2.5	95	-0.03
24	Anthracene	0.400	0.392	2.0	92	-0.01
25 SURR	Fluoranthene-d10	0.400	0.387	3.3	89	-0.02
26	Fluoranthene	0.400	0.389	2.8	90	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	87	-0.02
28	Pyrene	0.400	0.402	-0.5	90	-0.02
29 S	Terphenyl-d14	0.400	0.406	-1.5	89	-0.02
30	Benzo(a)anthracene	0.400	0.393	1.8	86	-0.02
31	Chrysene	0.400	0.413	-3.2	90	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.330	17.5	75	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.404	-1.0	86	-0.05
34 I	Perylene-d12	0.400	0.400	0.0	84	-0.03
35	Benzo(b)fluoranthene	0.400	0.399	0.3	87	-0.03
36	Benzo(k)fluoranthene	0.400	0.398	0.5	86	-0.03
37 C	Benzo(a)pyrene	0.400	0.385	3.8	83	-0.03
38	Dibenzo(a,h)anthracene	0.400	0.412	-3.0	88	-0.05
39	Benzo(g,h,i)perylene	0.400	0.407	-1.7	87	-0.06

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE040219\
Data File : BE099277.D
Acq On : 2 Apr 2019 16:56
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE040219

Quant Time: Apr 02 19:23:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 02 17:12:00 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				