

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE040219\
 Data File : BE099277.D
 Acq On : 2 Apr 2019 16:56
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE040219

Quant Time: Apr 02 19:23:34 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 02 17:12:00 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	-0.03
2	1,4-Dioxane	0.545	0.579	-6.2	104	-0.03
3	n-Nitrosodimethylamine	0.322	0.324	-0.6	102	-0.03
4 S	2-Fluorophenol	1.063	1.085	-2.1	95	-0.03
5 S	Phenol-d6	1.461	1.426	2.4	94	-0.01
6	bis(2-Chloroethyl)ether	1.310	1.319	-0.7	98	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	96	-0.01
8 S	Nitrobenzene-d5	0.275	0.289	-5.1	105	-0.02
9	Naphthalene	4.351	4.357	-0.1	96	-0.02
10	Hexachlorobutadiene	0.283	0.286	-1.1	96	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.683	2.1	95	-0.01
12	2-Methylnaphthalene	0.740	0.711	3.9	93	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	96	-0.03
14 S	2,4,6-Tribromophenol	0.184	0.156	15.2	91	-0.01
15 S	2-Fluorobiphenyl	1.572	1.546	1.7	95	-0.01
16	Acenaphthylene	1.895	1.736	8.4	91	-0.01
17	Acenaphthene	1.118	1.075	3.8	95	-0.01
18	Fluorene	1.615	1.519	5.9	90	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	-0.03
20	4-Bromophenyl-phenylether	0.227	0.220	3.1	93	-0.03
21	Hexachlorobenzene	0.216	0.221	-2.3	97	-0.03
22	Pentachlorophenol	0.090	0.073	18.9	87	-0.03
23	Phenanthrene	1.035	1.061	-2.5	95	-0.03
24	Anthracene	0.976	0.957	1.9	92	-0.01
25 SURR	Fluoranthene-d10	5.280	5.103	3.4	89	-0.02
26	Fluoranthene	1.528	1.487	2.7	90	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	87	-0.02
28	Pyrene	1.056	1.062	-0.6	90	-0.02
29 S	Terphenyl-d14	0.730	0.740	-1.4	89	-0.02
30	Benzo(a)anthracene	1.259	1.237	1.7	86	-0.02
31	Chrysene	1.235	1.274	-3.2	90	-0.01
32	Bis(2-ethylhexyl)phthalate	1.605	1.323	17.6	75	-0.02
33	Indeno(1,2,3-cd)pyrene	1.412	1.426	-1.0	86	-0.05
34 I	Perylene-d12	1.000	1.000	0.0	84	-0.03
35	Benzo(b)fluoranthene	1.194	1.191	0.3	87	-0.03
36	Benzo(k)fluoranthene	1.162	1.157	0.4	86	-0.03
37 C	Benzo(a)pyrene	1.122	1.079	3.8	83	-0.03
38	Dibenzo(a,h)anthracene	1.116	1.150	-3.0	88	-0.05
39	Benzo(g,h,i)perylene	1.112	1.131	-1.7	87	-0.06

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE040219\
Data File : BE099277.D
Acq On : 2 Apr 2019 16:56
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE040219

Quant Time: Apr 02 19:23:34 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 02 17:12:00 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		