

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040221\
 Data File : BE101189.D
 Acq On : 2 Apr 2021 8:36
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 02 10:05:10 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 02 10:03:59 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	96	0.00
2	1,4-Dioxane	0.400	0.469	-17.2	99	0.00
3	n-Nitrosodimethylamine	0.400	0.420	-5.0	93	0.00
4 S	2-Fluorophenol	0.400	0.505	-26.2#	107	0.00
5 S	Phenol-d6	0.400	0.489	-22.2	105	0.00
6	bis(2-Chloroethyl)ether	0.400	0.444	-11.0	95	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.475	-18.7	123	0.00
9	Naphthalene	0.400	0.417	-4.2	96	0.00
10	Hexachlorobutadiene	0.400	0.388	3.0	88	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.416	-4.0	97	0.00
12	2-Methylnaphthalene	0.400	0.417	-4.2	99	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	108	0.00
14 S	2,4,6-Tribromophenol	0.400	0.624	-56.0#	185	0.00
15 S	2-Fluorobiphenyl	0.400	0.387	3.3	96	0.00
16	Acenaphthylene	0.400	0.368	8.0	96	0.00
17	Acenaphthene	0.400	0.408	-2.0	106	0.00
18	Fluorene	0.400	0.401	-0.3	101	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.683	-70.8#	172	0.00
21	4-Bromophenyl-phenylether	0.400	0.426	-6.5	105	0.00
22	Hexachlorobenzene	0.400	0.408	-2.0	97	0.00
23	Atrazine	0.400	0.517	-29.3#	136	0.00
24	Pentachlorophenol	0.400	0.478	-19.5	150	0.00
25	Phenanthrene	0.400	0.409	-2.2	98	0.00
26	Anthracene	0.400	0.384	4.0	96	0.00
27 SURR	Fluoranthene-d10	0.400	0.389	2.8	93	0.00
28	Fluoranthene	0.400	0.382	4.5	90	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	98	0.00
30	Pyrene	0.400	0.423	-5.7	91	0.00
31 S	Terphenyl-d14	0.400	0.432	-8.0	93	0.00
32	Benzo(a)anthracene	0.400	0.455	-13.7	107	0.00
33	Chrysene	0.400	0.428	-7.0	90	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.574	-43.5#	149	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.582	-45.5#	139	0.00
36 I	Perylene-d12	0.400	0.400	0.0	112	0.00
37	Benzo(b)fluoranthene	0.400	0.441	-10.2	119	0.00
38	Benzo(k)fluoranthene	0.400	0.388	3.0	101	0.00
39 C	Benzo(a)pyrene	0.400	0.396	1.0	110	0.00
40	Dibenzo(a,h)anthracene	0.400	0.496	-24.0	144	0.00
41	Benzo(g,h,i)perylene	0.400	0.441	-10.2	115	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_E

LabSampleId :

SSTDCCC0.4