

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040519\
 Data File : BE099307.D
 Acq On : 5 Apr 2019 14:04
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 05 15:01:53 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 05 14:55:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	-0.03
2	1,4-Dioxane	0.545	0.552	-1.3	119	-0.01
3	n-Nitrosodimethylamine	0.322	0.308	4.3	116	-0.03
4 S	2-Fluorophenol	1.063	1.049	1.3	110	-0.03
5 S	Phenol-d6	1.461	1.424	2.5	113	-0.01
6	bis(2-Chloroethyl)ether	1.310	1.293	1.3	115	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	110	-0.01
8 S	Nitrobenzene-d5	0.275	0.315	-14.5	132	-0.03
9	Naphthalene	4.351	4.659	-7.1	118	-0.03
10	Hexachlorobutadiene	0.283	0.310	-9.5	120	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.728	-4.3	117	-0.03
12	2-Methylnaphthalene	0.740	0.793	-7.2	119	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	-0.03
14 S	2,4,6-Tribromophenol	0.184	0.180	2.2	119	-0.01
15 S	2-Fluorobiphenyl	1.572	1.709	-8.7	118	-0.01
16	Acenaphthylene	1.895	1.909	-0.7	114	-0.01
17	Acenaphthene	1.118	1.140	-2.0	114	-0.03
18	Fluorene	1.615	1.662	-2.9	112	-0.03
19 I	Phenanthrene-d10	1.000	1.000	0.0	107	-0.03
20	4-Bromophenyl-phenylether	0.227	0.243	-7.0	117	-0.03
21	Hexachlorobenzene	0.216	0.234	-8.3	118	-0.03
22	Pentachlorophenol	0.090	0.106	-17.8	145	-0.03
23	Phenanthrene	1.035	1.121	-8.3	115	-0.03
24	Anthracene	0.976	0.971	0.5	107	-0.01
25 SURR	Fluoranthene-d10	5.280	5.430	-2.8	109	-0.02
26	Fluoranthene	1.528	1.604	-5.0	112	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	99	-0.04
28	Pyrene	1.056	1.142	-8.1	109	-0.02
29 S	Terphenyl-d14	0.730	0.807	-10.5	111	-0.02
30	Benzo(a)anthracene	1.259	1.235	1.9	97	-0.02
31	Chrysene	1.235	1.340	-8.5	107	-0.01
32	Bis(2-ethylhexyl)phthalate	1.605	1.173	26.9#	75	-0.02
33	Indeno(1,2,3-cd)pyrene	1.412	1.485	-5.2	102	-0.07
34 I	Perylene-d12	1.000	1.000	0.0	96	-0.04
35	Benzo(b)fluoranthene	1.194	1.266	-6.0	105	-0.03
36	Benzo(k)fluoranthene	1.162	1.264	-8.8	106	-0.03
37 C	Benzo(a)pyrene	1.122	1.140	-1.6	100	-0.04
38	Dibenzo(a,h)anthracene	1.116	1.189	-6.5	104	-0.07
39	Benzo(g,h,i)perylene	1.112	1.164	-4.7	102	-0.07

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040519\
Data File : BE099307.D
Acq On : 5 Apr 2019 14:04
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 05 15:01:53 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 05 14:55:44 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			