

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040521\
 Data File : BE101199.D
 Acq On : 5 Apr 2021 17:10
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Quant Time: Apr 06 02:25:06 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 06 02:23:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.841	152	3754	0.40	ng	0.00
7) Naphthalene-d8	10.614	136	15844	0.40	ng	0.00
13) Acenaphthene-d10	14.458	164	9805	0.40	ng	0.00
19) Phenanthrene-d10	17.210	188	25034	0.40	ng	# 0.01
29) Chrysene-d12	21.367	240	32081	0.40	ng	# 0.01
36) Perylene-d12	23.838	264	31194	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.422	112	1164	0.13	ng	0.00
5) Phenol-d6	6.989	99	1621	0.12	ng	0.00
8) Nitrobenzene-d5	8.977	82	1228	0.13	ng	0.00
11) 2-Methylnaphthalene-d10	12.196	152	3528	0.14	ng	0.00
14) 2,4,6-Tribromophenol	15.956	330	278	0.16	ng	0.00
15) 2-Fluorobiphenyl	13.075	172	3723	0.10	ng	0.00
27) Fluoranthene-d10	19.220	212	42261	0.13	ng	0.00
31) Terphenyl-d14	19.818	244	7214	0.10	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.372	88	877	0.14	ng	89
3) n-Nitrosodimethylamine	3.683	42	654	0.10	ng	# 81
6) bis(2-Chloroethyl)ether	7.254	93	1305	0.11	ng	99
9) Naphthalene	10.666	128	17331	0.11	ng	99
10) Hexachlorobutadiene	10.964	225	812	0.11	ng	94
12) 2-Methylnaphthalene	12.268	142	2924	0.11	ng	96
16) Acenaphthylene	14.170	152	3599	0.10	ng	99
17) Acenaphthene	14.516	154	2832	0.10	ng	98
18) Fluorene	15.517	166	3653	0.10	ng	100
20) 4,6-Dinitro-2-methylph...	15.578	198	224	0.18	ng	# 18
21) 4-Bromophenyl-phenylether	16.409	248	1243	0.11	ng	# 92
22) Hexachlorobenzene	16.515	284	1279	0.10	ng	96
23) Atrazine	16.666	200	842	0.12	ng	# 84
24) Pentachlorophenol	16.863	266	409	0.13	ng	94
25) Phenanthrene	17.241	178	7478	0.11	ng	98
26) Anthracene	17.331	178	5899	0.10	ng	98
28) Fluoranthene	19.249	202	8823	0.10	ng	99
30) Pyrene	19.611	202	9351	0.10	ng	100
32) Benzo(a)anthracene	21.343	228	9844	0.12	ng	99
33) Chrysene	21.403	228	10880	0.11	ng	100
34) Bis(2-ethylhexyl)phtha...	21.282	149	14095	0.26	ng	100
35) Indeno(1,2,3-cd)pyrene	26.456	276	9342	0.14	ng	99
37) Benzo(b)fluoranthene	23.079	252	9145	0.11	ng	# 96
38) Benzo(k)fluoranthene	23.079	252	9145	0.09	ng	98
39) Benzo(a)pyrene	23.728	252	7445	0.10	ng	# 91
40) Dibenzo(a,h)anthracene	26.484	278	7702	0.12	ng	96
41) Benzo(g,h,i)perylene	27.254	276	9322	0.12	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040521\
Data File : BE101199.D
Acq On : 5 Apr 2021 17:10
Operator : CG/JU
Sample : SSTDIC0.1
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Quant Time: Apr 06 02:25:06 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 06 02:23:26 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.841 min Scan# 432
 Delta R.T. 0.002 min
 Lab File: BE101199.D
 Acq: 5 Apr 2021 17:10




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 3754
 Ion Ratio Lower Upper
 152 100
 150 150.9 119.2 178.8
 115 58.8 46.6 70.0



#2
 1,4-Dioxane
 Concen: 0.14 ng
 RT: 3.372 min Scan# 44
 Delta R.T. 0.002 min
 Lab File: BE101199.D
 Acq: 5 Apr 2021 17:10

Tgt Ion: 88 Resp: 877
 Ion Ratio Lower Upper
 88 100
 58 71.0 66.1 99.1
 43 31.6 28.8 43.2









#3
n-Nitrosodimethylamine
Concen: 0.10 ng
RT: 3.683 min Scan# 71
Delta R.T. 0.002 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion: 42 Resp: 654
Ion Ratio Lower Upper
42 100
74 131.8 89.7 134.5
44 0.0 5.4 8.0#



#4
2-Fluorophenol
Concen: 0.13 ng
RT: 5.422 min Scan# 222
Delta R.T. 0.002 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion: 112 Resp: 1164
Ion Ratio Lower Upper
112 100
64 76.7 61.5 92.3
63 155.1 120.2 180.4









#5
Phenol-d6
Concen: 0.12 ng
RT: 6.989 min Scan# 358
Delta R.T. -0.010 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion: 99 Resp: 1621
Ion Ratio Lower Upper
99 100
42 32.4 18.7 28.1#
71 34.5 26.6 40.0



#6
bis(2-Chloroethyl)ether
Concen: 0.11 ng
RT: 7.254 min Scan# 381
Delta R.T. -0.010 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion: 93 Resp: 1305
Ion Ratio Lower Upper
93 100
63 372.5 295.8 443.6
95 34.6 25.8 38.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.614 min Scan# 650
Delta R.T. -0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion:	136	Resp:	15844
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.0	13.4
54	56.2	40.3	60.5
68	10.3	7.9	11.9



#8
Nitrobenzene-d5
Concen: 0.13 ng
RT: 8.977 min Scan# 524
Delta R.T. -0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:	82	Resp:	1228
Ion	Ratio	Lower	Upper
82	100		
128	145.6	110.4	165.6
54	238.4	170.7	256.1









#9
Naphthalene
Concen: 0.11 ng
RT: 10.666 min Scan# 654
Delta R.T. -0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion:	128	Resp:	17331
Ion	Ratio	Lower	Upper
128	100		
129	3.2	2.2	3.4
127	3.5	2.6	4.0



#10
Hexachlorobutadiene
Concen: 0.11 ng
RT: 10.964 min Scan# 677
Delta R.T. -0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:	225	Resp:	812
Ion	Ratio	Lower	Upper
225	100		
223	61.9	51.8	77.6
227	58.9	52.9	79.3









#11
2-Methylnaphthalene-d10
Concen: 0.14 ng
RT: 12.196 min Scan# 763
Delta R.T. 0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 3528
Ion Ratio Lower Upper
152 100
151 20.5 16.6 24.8



#12
2-Methylnaphthalene
Concen: 0.11 ng
RT: 12.268 min Scan# 768
Delta R.T. 0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:142 Resp: 2924
Ion Ratio Lower Upper
142 100
141 92.9 71.9 107.9
115 38.1 27.7 41.5









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.458 min Scan# 920
Delta R.T. 0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion:164	Resp:	9805
Ion Ratio	Lower	Upper
164	100	
162	99.9	80.0 120.0
160	44.0	34.1 51.1



#14
2,4,6-Tribromophenol
Concen: 0.16 ng
RT: 15.956 min Scan# 1016
Delta R.T. -0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:330	Resp:	278
Ion Ratio	Lower	Upper
330	100	
332	100.7	75.8 113.8
141	41.7	30.5 45.7









#15
2-Fluorobiphenyl
Concen: 0.10 ng
RT: 13.075 min Scan# 824
Delta R.T. 0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion:172 Resp: 3723




Ion	Ratio	Lower	Upper
172	100		
171	35.5	28.4	42.6
170	26.4	19.5	29.3




#16
Acenaphthylene
Concen: 0.10 ng
RT: 14.170 min Scan# 900
Delta R.T. 0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:152 Resp: 3599

Ion	Ratio	Lower	Upper
152	100		
151	19.0	15.3	22.9
153	14.1	10.3	15.5





#17

Acenaphthene

Concen: 0.10 ng

RT: 14.516 min Scan# 924

Delta R.T. 0.001 min

Lab File: BE101199.D

Acq: 5 Apr 2021 17:10

Instrument :

BNA_E

ClientSampleId :

SSTDIC0.1





Tgt Ion:154 Resp: 2832

Ion Ratio Lower Upper

154 100

153 108.9 86.7 130.1

152 56.6 42.6 64.0



#18

Fluorene

Concen: 0.10 ng

RT: 15.517 min Scan# 987

Delta R.T. 0.014 min

Lab File: BE101199.D

Acq: 5 Apr 2021 17:10




Tgt Ion:166 Resp: 3653


Ion Ratio Lower Upper

166 100

165 98.8 79.1 118.7

167 14.2 11.0 16.4









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.210 min Scan# 1099
Delta R.T. 0.014 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion:188 Resp: 25034
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.1 12.9 19.3#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.18 ng
RT: 15.578 min Scan# 991
Delta R.T. -0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:198 Resp: 224
Ion Ratio Lower Upper
198 100
51 153.3 59.0 88.4#
105 77.2 31.9 47.9#









#21
4-Bromophenyl-phenylether
Concen: 0.11 ng
RT: 16.409 min Scan# 1046
Delta R.T. -0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1


Tgt Ion:248 Resp: 1243
Ion Ratio Lower Upper
248 100
250 101.4 85.3 127.9
141 33.2 19.0 28.4#



#22
Hexachlorobenzene
Concen: 0.10 ng
RT: 16.515 min Scan# 1053
Delta R.T. -0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:284 Resp: 1279
Ion Ratio Lower Upper
284 100
142 34.3 26.8 40.2
249 36.4 26.5 39.7





#23

Atrazine

Concen: 0.12 ng

RT: 16.666 min Scan# 1063

Delta R.T. -0.001 min

Lab File: BE101199.D

Acq: 5 Apr 2021 17:10





Instrument :

BNA_E

ClientSampleId :

SSTDIC0.1

Tgt Ion:	200	Resp:	842
Ion Ratio	Lower	Upper	
200	100		
173	38.5	19.8	29.8#
215	55.6	38.7	58.1



#24

Pentachlorophenol

Concen: 0.13 ng




RT: 16.863 min Scan# 1076


Delta R.T. 0.014 min

Lab File: BE101199.D

Acq: 5 Apr 2021 17:10

Tgt Ion:	266	Resp:	409
Ion Ratio	Lower	Upper	
266	100		
264	61.4	48.9	73.3
268	71.1	49.8	74.6









#25
Phenanthrene
Concen: 0.11 ng
RT: 17.241 min Scan# 1101
Delta R.T. -0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion:178 Resp: 7478
Ion Ratio Lower Upper
178 100
176 19.1 14.8 22.2
179 16.3 12.1 18.1



#26
Anthracene
Concen: 0.10 ng
RT: 17.331 min Scan# 1107
Delta R.T. -0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:178 Resp: 5899
Ion Ratio Lower Upper
178 100
176 19.4 14.6 22.0
179 15.3 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.13 ng
RT: 19.220 min Scan# 1359
Delta R.T. 0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion:212 Resp: 42261
Ion Ratio Lower Upper
212 100
106 3.2 2.7 4.1
104 1.9 1.4 2.2



#28
Fluoranthene
Concen: 0.10 ng
RT: 19.249 min Scan# 1364
Delta R.T. 0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:202 Resp: 8823
Ion Ratio Lower Upper
202 100
101 12.0 9.2 13.8
203 16.5 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.367 min Scan# 1605
Delta R.T. 0.013 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1


Tgt Ion:240 Resp: 32081
Ion Ratio Lower Upper
240 100
120 4.2 10.5 15.7#
236 24.9 21.8 32.8



#30
Pyrene
Concen: 0.10 ng
RT: 19.611 min Scan# 1427
Delta R.T. 0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:202 Resp: 9351
Ion Ratio Lower Upper
202 100
200 20.8 16.6 25.0
203 17.6 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.10 ng
 RT: 19.818 min Scan# 1463
 Delta R.T. 0.001 min
 Lab File: BE101199.D
 Acq: 5 Apr 2021 17:10




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244	Resp:	7214
Ion Ratio	Lower	Upper
244	100	
212	30.4	22.4 33.6
122	11.4	7.8 11.8



#32
 Benzo(a)anthracene
 Concen: 0.12 ng
 RT: 21.343 min Scan# 1603
 Delta R.T. 0.001 min
 Lab File: BE101199.D
 Acq: 5 Apr 2021 17:10

Tgt Ion:228	Resp:	9844
Ion Ratio	Lower	Upper
228	100	
226	27.0	21.5 32.3
229	20.5	15.9 23.9









#33
Chrysene
Concen: 0.11 ng
RT: 21.403 min Scan# 1608
Delta R.T. 0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion:228	Resp:	10880
Ion Ratio	Lower	Upper
228	100	
226	28.6	22.9 34.3
229	20.6	16.2 24.2



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.26 ng
RT: 21.282 min Scan# 1598
Delta R.T. 0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:149	Resp:	14095
Ion Ratio	Lower	Upper
149	100	
167	7.1	5.8 8.6
279	1.2	0.9 1.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.14 ng
RT: 26.456 min Scan# 2896
Delta R.T. 0.003 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Tgt Ion:276 Resp: 9342
Ion Ratio Lower Upper
276 100
138 21.5 17.8 26.6
277 24.4 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.838 min Scan# 2166
Delta R.T. 0.001 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:264 Resp: 31194
Ion Ratio Lower Upper
264 100
260 25.4 19.8 29.8
265 26.3 21.4 32.0





#37

Benzo(b)fluoranthene

Concen: 0.11 ng

RT: 23.079 min Scan# 1953

Delta R.T. 0.001 min

Lab File: BE101199.D

Acq: 5 Apr 2021 17:10





Tgt Ion:252 Resp: 9145

Ion Ratio Lower Upper

252 100

253 23.9 18.2 27.2

125 12.2 7.7 11.5#



#38

Benzo(k)fluoranthene

Concen: 0.09 ng

RT: 23.079 min Scan# 1953

Delta R.T. -0.049 min

Lab File: BE101199.D

Acq: 5 Apr 2021 17:10




Tgt Ion:252 Resp: 9145


Ion Ratio Lower Upper

252 100

253 23.9 18.2 27.2

125 12.2 9.6 14.4









#39
Benzo(a)pyrene
Concen: 0.10 ng
RT: 23.728 min Scan# 2135
Delta R.T. 0.005 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10




Instrument :
BNA_E
ClientSampleId :
SSTDICC0.1


Tgt Ion:252 Resp: 7445
Ion Ratio Lower Upper
252 100
253 26.8 17.8 26.8#
125 14.4 8.7 13.1#



#40
Dibenzo(a,h)anthracene
Concen: 0.12 ng
RT: 26.484 min Scan# 2904
Delta R.T. 0.010 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Tgt Ion:278 Resp: 7702
Ion Ratio Lower Upper
278 100
139 16.8 12.3 18.5
279 26.2 19.5 29.3





#41
Benzo(g,h,i)perylene
Concen: 0.12 ng
RT: 27.254 min Scan# 3120
Delta R.T. 0.003 min
Lab File: BE101199.D
Acq: 5 Apr 2021 17:10

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion: 276 Resp: 9322
Ion Ratio Lower Upper
276 100
277 24.7 18.7 28.1
138 20.0 15.0 22.6

