

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040521\
 Data File : BE101206.D
 Acq On : 5 Apr 2021 22:05
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE040521

Quant Time: Apr 06 08:05:47 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 06 03:03:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.710	0.742	-4.5	97	0.00
3	n-Nitrosodimethylamine	0.723	0.738	-2.1	98	0.00
4 S	2-Fluorophenol	1.089	1.108	-1.7	98	0.00
5 S	Phenol-d6	1.605	1.627	-1.4	99	0.00
6	bis(2-Chloroethyl)ether	1.313	1.373	-4.6	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.315	0.298	5.4	98	0.00
9	Naphthalene	4.332	4.443	-2.6	99	0.00
10	Hexachlorobutadiene	0.192	0.192	0.0	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.894	0.883	1.2	98	0.00
12	2-Methylnaphthalene	0.750	0.746	0.5	98	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	-0.01
14 S	2,4,6-Tribromophenol	0.125	0.103	17.6	93	0.00
15 S	2-Fluorobiphenyl	1.455	1.479	-1.6	100	0.00
16	Acenaphthylene	1.592	1.504	5.5	102	0.00
17	Acenaphthene	1.163	1.176	-1.1	104	0.00
18	Fluorene	1.499	1.478	1.4	104	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
20	4,6-Dinitro-2-methylphenol	0.052	0.041	21.2	109	0.00
21	4-Bromophenyl-phenylether	0.209	0.193	7.7	97	0.00
22	Hexachlorobenzene	0.211	0.212	-0.5	103	0.00
23	Atrazine	0.161	0.147	8.7	106	0.00
24	Pentachlorophenol	0.083	0.061	26.5#	99	0.00
25	Phenanthrene	1.183	1.208	-2.1	106	0.00
26	Anthracene	1.020	0.995	2.5	109	0.00
27 SURR	Fluoranthene-d10	6.882	6.710	2.5	106	0.00
28	Fluoranthene	1.505	1.467	2.5	104	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
30	Pyrene	1.164	1.157	0.6	106	0.00
31 S	Terphenyl-d14	0.903	0.948	-5.0	106	0.00
32	Benzo(a)anthracene	1.208	1.200	0.7	105	0.00
33	Chrysene	1.330	1.339	-0.7	103	0.00
34	Bis(2-ethylhexyl)phthalate	1.492	1.274	14.6	117	0.00
35	Indeno(1,2,3-cd)pyrene	1.169	1.153	1.4	103	0.00
36 I	Perylene-d12	1.000	1.000	0.0	103	0.00
37	Benzo(b)fluoranthene	1.227	1.166	5.0	101	0.00
38	Benzo(k)fluoranthene	1.384	1.386	-0.1	107	0.00
39 C	Benzo(a)pyrene	1.086	1.060	2.4	102	0.00
40	Dibenzo(a,h)anthracene	1.062	1.003	5.6	102	0.00
41	Benzo(g,h,i)perylene	1.150	1.111	3.4	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_E

ClientSampleId :

ICVBE040521