

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040521\
 Data File : BE101206.D
 Acq On : 5 Apr 2021 22:05
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE040521

Quant Time: Apr 06 08:05:47 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 06 03:03:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.839	152	3934	0.40	ng	0.00
7) Naphthalene-d8	10.614	136	16590	0.40	ng	0.00
13) Acenaphthene-d10	14.442	164	10634	0.40	ng	-0.01
19) Phenanthrene-d10	17.196	188	25665	0.40	ng	0.00
29) Chrysene-d12	21.354	240	32615	0.40	ng	0.00
36) Perylene-d12	23.838	264	31742	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.421	112	4360	0.41	ng	0.00
5) Phenol-d6	6.999	99	6401	0.41	ng	0.00
8) Nitrobenzene-d5	8.977	82	4952	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.195	152	14649	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.957	330	1091	0.33	ng	0.00
15) 2-Fluorobiphenyl	13.074	172	15731	0.41	ng	0.00
27) Fluoranthene-d10	19.220	212	172215	0.39	ng	0.00
31) Terphenyl-d14	19.817	244	30923	0.42	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.371	88	2919	0.42	ng	97
3) n-Nitrosodimethylamine	3.682	42	2903	0.41	ng	# 96
6) bis(2-Chloroethyl)ether	7.263	93	5402	0.42	ng	98
9) Naphthalene	10.666	128	73708	0.41	ng	100
10) Hexachlorobutadiene	10.965	225	3192	0.40	ng	98
12) 2-Methylnaphthalene	12.267	142	12371	0.40	ng	99
16) Acenaphthylene	14.169	152	15997	0.38	ng	100
17) Acenaphthene	14.514	154	12501	0.40	ng	99
18) Fluorene	15.503	166	15717	0.39	ng	100
20) 4,6-Dinitro-2-methylph...	15.579	198	1040	0.43	ng	93
21) 4-Bromophenyl-phenylether	16.410	248	4955	0.37	ng	97
22) Hexachlorobenzene	16.516	284	5446	0.40	ng	100
23) Atrazine	16.667	200	3762	0.36	ng	98
24) Pentachlorophenol	16.849	266	1574	0.30	ng	98
25) Phenanthrene	17.242	178	31016	0.41	ng	100
26) Anthracene	17.332	178	25536	0.39	ng	99
28) Fluoranthene	19.248	202	37644	0.39	ng	99
30) Pyrene	19.610	202	37720	0.40	ng	100
32) Benzo(a)anthracene	21.342	228	39129	0.40	ng	99
33) Chrysene	21.403	228	43657	0.40	ng	100
34) Bis(2-ethylhexyl)phtha...	21.282	149	41539	0.34	ng	100
35) Indeno(1,2,3-cd)pyrene	26.448	276	37595	0.39	ng	99
37) Benzo(b)fluoranthene	23.075	252	37014	0.38	ng	99
38) Benzo(k)fluoranthene	23.125	252	43985	0.40	ng	98
39) Benzo(a)pyrene	23.727	252	33637	0.39	ng	100
40) Dibenzo(a,h)anthracene	26.476	278	31833	0.38	ng	99
41) Benzo(g,h,i)perylene	27.250	276	35253	0.39	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040521\
Data File : BE101206.D
Acq On : 5 Apr 2021 22:05
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE040521

Quant Time: Apr 06 08:05:47 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 06 03:03:09 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.839 min Scan# 432
 Delta R.T. 0.000 min
 Lab File: BE101206.D
 Acq: 5 Apr 2021 22:05




Instrument :
 BNA_E
 ClientSampleId :
 ICVBE040521


Tgt Ion:152 Resp: 3934
 Ion Ratio Lower Upper
 152 100
 150 153.5 119.2 178.8
 115 57.6 46.6 70.0



#2
 1,4-Dioxane
 Concen: 0.42 ng
 RT: 3.371 min Scan# 44
 Delta R.T. -0.000 min
 Lab File: BE101206.D
 Acq: 5 Apr 2021 22:05

Tgt Ion: 88 Resp: 2919
 Ion Ratio Lower Upper
 88 100
 58 83.7 66.1 99.1
 43 32.5 28.8 43.2









#3
n-Nitrosodimethylamine
Concen: 0.41 ng
RT: 3.682 min Scan# 71
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion: 42 Resp: 2903
Ion Ratio Lower Upper
42 100
74 115.6 89.7 134.5
44 13.9 5.4 8.0#



#4
2-Fluorophenol
Concen: 0.41 ng
RT: 5.421 min Scan# 222
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion: 112 Resp: 4360
Ion Ratio Lower Upper
112 100
64 78.8 61.5 92.3
63 152.9 120.2 180.4









#5
Phenol-d6
Concen: 0.41 ng
RT: 6.999 min Scan# 359
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Instrument :
BNA_E
ClientSampleId :
ICVBE040521

Tgt Ion: 99 Resp: 6401




Ion	Ratio	Lower	Upper
99	100		
42	23.8	18.7	28.1
71	33.2	26.6	40.0




#6
bis(2-Chloroethyl)ether
Concen: 0.42 ng
RT: 7.263 min Scan# 382
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion: 93 Resp: 5402

Ion	Ratio	Lower	Upper
93	100		
63	366.1	295.8	443.6
95	32.5	25.8	38.6









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.614 min Scan# 650
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:	136	Resp:	16590
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.4
54	46.3	40.3	60.5
68	9.5	7.9	11.9



#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.977 min Scan# 524
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:	82	Resp:	4952
Ion	Ratio	Lower	Upper
82	100		
128	141.2	110.4	165.6
54	229.8	170.7	256.1









#9
Naphthalene
Concen: 0.41 ng
RT: 10.666 min Scan# 654
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:	128	Resp:	73708
Ion	Ratio	Lower	Upper
128	100		
129	2.9	2.2	3.4
127	3.2	2.6	4.0



#10
Hexachlorobutadiene
Concen: 0.40 ng
RT: 10.965 min Scan# 677
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:	225	Resp:	3192
Ion	Ratio	Lower	Upper
225	100		
223	63.3	51.8	77.6
227	64.1	52.9	79.3









#11
2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.195 min Scan# 763
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:152 Resp: 14649
Ion Ratio Lower Upper
152 100
151 21.0 16.6 24.8



#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.267 min Scan# 768
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:142 Resp: 12371
Ion Ratio Lower Upper
142 100
141 91.2 71.9 107.9
115 34.1 27.7 41.5









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.442 min Scan# 919
Delta R.T. -0.014 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:164 Resp: 10634
Ion Ratio Lower Upper
164 100
162 108.1 80.0 120.0
160 49.6 34.1 51.1



#14
2,4,6-Tribromophenol
Concen: 0.33 ng
RT: 15.957 min Scan# 1016
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:330 Resp: 1091
Ion Ratio Lower Upper
330 100
332 97.1 75.8 113.8
141 38.9 30.5 45.7









#15
2-Fluorobiphenyl
Concen: 0.41 ng
RT: 13.074 min Scan# 824
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Instrument :
BNA_E
ClientSampleId :
ICVBE040521

Tgt Ion:172 Resp: 15731




Ion	Ratio	Lower	Upper
172	100		
171	36.5	28.4	42.6
170	25.1	19.5	29.3




#16
Acenaphthylene
Concen: 0.38 ng
RT: 14.169 min Scan# 900
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:152 Resp: 15997

Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.3	22.9
153	12.5	10.3	15.5









#17
Acenaphthene
Concen: 0.40 ng
RT: 14.514 min Scan# 924
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Instrument :
BNA_E
ClientSampleId :
ICVBE040521

Tgt Ion:154 Resp: 12501




Ion	Ratio	Lower	Upper
154	100		
153	107.1	86.7	130.1
152	53.0	42.6	64.0




#18
Fluorene
Concen: 0.39 ng
RT: 15.503 min Scan# 986
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:166 Resp: 15717

Ion	Ratio	Lower	Upper
166	100		
165	98.9	79.1	118.7
167	13.6	11.0	16.4









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.196 min Scan# 1098
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:188 Resp: 25665
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 17.3 12.9 19.3



#20
4,6-Dinitro-2-methylphenol
Concen: 0.43 ng
RT: 15.579 min Scan# 991
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:198 Resp: 1040
Ion Ratio Lower Upper
198 100
51 67.5 59.0 88.4
105 35.1 31.9 47.9









#21
4-Bromophenyl-phenylether
Concen: 0.37 ng
RT: 16.410 min Scan# 1046
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:248 Resp: 4955
Ion Ratio Lower Upper
248 100
250 110.3 85.3 127.9
141 23.1 19.0 28.4



#22
Hexachlorobenzene
Concen: 0.40 ng
RT: 16.516 min Scan# 1053
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:284 Resp: 5446
Ion Ratio Lower Upper
284 100
142 33.3 26.8 40.2
249 33.4 26.5 39.7









#23
Atrazine
Concen: 0.36 ng
RT: 16.667 min Scan# 1063
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:200 Resp: 3762
Ion Ratio Lower Upper
200 100
173 22.7 19.8 29.8
215 48.5 38.7 58.1



#24
Pentachlorophenol
Concen: 0.30 ng
RT: 16.849 min Scan# 1075
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:266 Resp: 1574
Ion Ratio Lower Upper
266 100
264 60.7 48.9 73.3
268 64.3 49.8 74.6









#25
Phenanthrene
Concen: 0.41 ng
RT: 17.242 min Scan# 1101
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampled :
ICVBE040521


Tgt Ion:178 Resp: 31016
Ion Ratio Lower Upper
178 100
176 18.6 14.8 22.2
179 15.3 12.1 18.1



#26
Anthracene
Concen: 0.39 ng
RT: 17.332 min Scan# 1107
Delta R.T. -0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:178 Resp: 25536
Ion Ratio Lower Upper
178 100
176 18.8 14.6 22.0
179 15.2 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.39 ng
RT: 19.220 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:212 Resp: 172215
Ion Ratio Lower Upper
212 100
106 3.4 2.7 4.1
104 1.9 1.4 2.2



#28
Fluoranthene
Concen: 0.39 ng
RT: 19.248 min Scan# 1364
Delta R.T. 0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:202 Resp: 37644
Ion Ratio Lower Upper
202 100
101 11.6 9.2 13.8
203 17.6 13.8 20.6









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.354 min Scan# 1604
Delta R.T. 0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:240 Resp: 32615
Ion Ratio Lower Upper
240 100
120 12.5 10.5 15.7
236 27.5 21.8 32.8



#30
Pyrene
Concen: 0.40 ng
RT: 19.610 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:202 Resp: 37720
Ion Ratio Lower Upper
202 100
200 20.8 16.6 25.0
203 17.6 14.0 21.0









#31
Terphenyl-d14
Concen: 0.42 ng
RT: 19.817 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:244 Resp: 30923
Ion Ratio Lower Upper
244 100
212 27.5 22.4 33.6
122 9.4 7.8 11.8



#32
Benzo(a)anthracene
Concen: 0.40 ng
RT: 21.342 min Scan# 1603
Delta R.T. 0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:228 Resp: 39129
Ion Ratio Lower Upper
228 100
226 26.8 21.5 32.3
229 19.4 15.9 23.9









#33
Chrysene
Concen: 0.40 ng
RT: 21.403 min Scan# 1608
Delta R.T. 0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:228 Resp: 43657
Ion Ratio Lower Upper
228 100
226 28.4 22.9 34.3
229 20.2 16.2 24.2



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.34 ng
RT: 21.282 min Scan# 1598
Delta R.T. 0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:149 Resp: 41539
Ion Ratio Lower Upper
149 100
167 7.2 5.8 8.6
279 1.2 0.9 1.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.39 ng
RT: 26.448 min Scan# 2894
Delta R.T. -0.005 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Instrument :
BNA_E
ClientSampleId :
ICVBE040521


Tgt Ion:276 Resp: 37595
Ion Ratio Lower Upper
276 100
138 21.3 17.8 26.6
277 24.7 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.838 min Scan# 2166
Delta R.T. 0.000 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Tgt Ion:264 Resp: 31742
Ion Ratio Lower Upper
264 100
260 24.2 19.8 29.8
265 25.6 21.4 32.0









#37
Benzo(b)fluoranthene
Concen: 0.38 ng
RT: 23.075 min Scan# 1952
Delta R.T. -0.003 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05




Tgt Ion:252 Resp: 37014
Ion Ratio Lower Upper
252 100
253 22.3 18.2 27.2
125 9.7 7.7 11.5


Instrument :
BNA_E
ClientSampleId :
ICVBE040521



#38
Benzo(k)fluoranthene
Concen: 0.40 ng
RT: 23.125 min Scan# 1966
Delta R.T. -0.003 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05


Tgt Ion:252 Resp: 43985
Ion Ratio Lower Upper
252 100
253 23.2 18.2 27.2
125 13.8 9.6 14.4








#39
Benzo(a)pyrene
Concen: 0.39 ng
RT: 23.727 min Scan# 2135
Delta R.T. 0.004 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05


Instrument :
BNA_E
ClientSampleId :
ICVBE040521





Tgt Ion:252 Resp: 33637
Ion Ratio Lower Upper
252 100
253 22.5 17.8 26.8
125 10.8 8.7 13.1




#40
Dibenzo(a,h)anthracene
Concen: 0.38 ng
RT: 26.476 min Scan# 2902
Delta R.T. 0.002 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05



Tgt Ion:278 Resp: 31833
Ion Ratio Lower Upper
278 100
139 14.5 12.3 18.5
279 24.5 19.5 29.3





#41
Benzo(g,h,i)perylene
Concen: 0.39 ng
RT: 27.250 min Scan# 3119
Delta R.T. -0.001 min
Lab File: BE101206.D
Acq: 5 Apr 2021 22:05

Instrument :
BNA_E
ClientSampleId :
ICVBE040521

Tgt Ion:276 Resp: 35253
Ion Ratio Lower Upper
276 100
277 24.5 18.7 28.1
138 20.1 15.0 22.6

