

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040721\
 Data File : BE101237.D
 Acq On : 7 Apr 2021 20:26
 Operator : CG/JU
 Sample : M1931-05DL 2X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MW-5DL

Quant Time: Apr 08 03:24:03 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 07 10:22:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.830	152	3871	0.40	ng	# 0.00
7) Naphthalene-d8	10.614	136	19991	0.40	ng	0.00
13) Acenaphthene-d10	14.443	164	16345	0.40	ng	0.00
19) Phenanthrene-d10	17.196	188	42796	0.40	ng	#-0.01
29) Chrysene-d12	21.353	240	45747	0.40	ng	0.00
36) Perylene-d12	23.830	264	32954	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.423	112	787	0.09	ng	0.00
5) Phenol-d6	6.989	99	752	0.05	ng	0.00
8) Nitrobenzene-d5	8.977	82	2831	0.19	ng	0.00
11) 2-Methylnaphthalene-d10	12.195	152	4960	0.11	ng	0.00
14) 2,4,6-Tribromophenol	15.956	330	1207	0.28	ng	0.00
15) 2-Fluorobiphenyl	13.074	172	7174	0.12	ng	0.00
27) Fluoranthene-d10	19.215	212	114345	0.15	ng	0.00
31) Terphenyl-d14	19.813	244	26597	0.22	ng	0.00
Target Compounds						
						Qvalue
9) Naphthalene	10.666	128	314880	1.45	ng	# 90
12) 2-Methylnaphthalene	12.267	142	138545	3.55	ng	98
16) Acenaphthylene	14.169	152	2643	0.04	ng	# 1
17) Acenaphthene	14.515	154	6323	0.13	ng	# 69
18) Fluorene	15.503	166	14883	0.23	ng	# 90
25) Phenanthrene	17.241	178	15737	0.12	ng	# 91
26) Anthracene	17.332	178	7314	0.07	ng	# 86
28) Fluoranthene	19.244	202	3803	0.02	ng	# 95
30) Pyrene	19.606	202	6580	0.04	ng	# 88
34) Bis(2-ethylhexyl)phtha...	21.281	149	19348	0.21	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040721\
Data File : BE101237.D
Acq On : 7 Apr 2021 20:26
Operator : CG/JU
Sample : M1931-05DL 2X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MW-5DL

Quant Time: Apr 08 03:24:03 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040621.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 07 10:22:45 2021
Response via : Initial Calibration

