

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE041219\
 Data File : BE099316.D
 Acq On : 12 Apr 2019 18:29
 Operator : JU/SJ
 Sample : K2366-10
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SA-SW124-20190409

Quant Time: Apr 12 20:33:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 02 17:16:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	4170	0.40	ng	-0.02
7) Naphthalene-d8	10.61	136	15327	0.40	ng	-0.01
13) Acenaphthene-d10	14.46	164	10178	0.40	ng	-0.03
19) Phenanthrene-d10	17.19	188	32211	0.40	ng	-0.03
27) Chrysene-d12	21.37	240	44886	0.40	ng	-0.02
34) Perylene-d12	23.85	264	44395	0.40	ng	-0.04
System Monitoring Compounds						
4) 2-Fluorophenol	5.40	112	402	0.04	ng	-0.01
5) Phenol-d6	6.98	99	129	0.01	ng	-0.01
8) Nitrobenzene-d5	8.96	82	2482	0.24	ng	-0.03
11) 2-Methylnaphthalene-d10	12.19	152	7956	0.30	ng	-0.03
14) 2,4,6-Tribromophenol	15.94	330	333	0.07	ng	-0.01
15) 2-Fluorobiphenyl	13.09	172	15286	0.38	ng	-0.01
25) Fluoranthene-d10	19.21	212	142667	0.34	ng	-0.03
29) Terphenyl-d14	19.81	244	33466	0.41	ng	-0.02
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.27	149	23793	0.132	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE041219\
Data File : BE099316.D
Acq On : 12 Apr 2019 18:29
Operator : JU/SJ
Sample : K2366-10
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SA-SW124-20190409

Quant Time: Apr 12 20:33:48 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 02 17:16:42 2019
Response via : Initial Calibration

