

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE041219\
 Data File : BE099309.D
 Acq On : 12 Apr 2019 14:01
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 12 17:55:14 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 12 17:51:18 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	123	0.00
2	1,4-Dioxane	0.400	0.417	-4.2	129	0.00
3	n-Nitrosodimethylamine	0.400	0.351	12.3	112	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	108	0.00
5 S	Phenol-d6	0.400	0.396	1.0	121	0.00
6	bis(2-Chloroethyl)ether	0.400	0.443	-10.7	136	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	124	0.00
8 S	Nitrobenzene-d5	0.400	0.392	2.0	127	0.00
9	Naphthalene	0.400	0.424	-6.0	131	0.00
10	Hexachlorobutadiene	0.400	0.416	-4.0	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.422	-5.5	133	0.00
12	2-Methylnaphthalene	0.400	0.438	-9.5	138	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.400	0.311	22.3	109	0.00
15 S	2-Fluorobiphenyl	0.400	0.434	-8.5	136	0.00
16	Acenaphthylene	0.400	0.385	3.8	125	0.00
17	Acenaphthene	0.400	0.423	-5.7	136	0.00
18	Fluorene	0.400	0.418	-4.5	130	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	129	0.00
20	4-Bromophenyl-phenylether	0.400	0.392	2.0	129	0.00
21	Hexachlorobenzene	0.400	0.397	0.8	129	0.00
22	Pentachlorophenol	0.400	0.395	1.3	146	0.00
23	Phenanthrene	0.400	0.428	-7.0	136	0.00
24	Anthracene	0.400	0.402	-0.5	130	0.00
25 SURR	Fluoranthene-d10	0.400	0.402	-0.5	127	0.00
26	Fluoranthene	0.400	0.413	-3.2	132	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	114	0.00
28	Pyrene	0.400	0.447	-11.7	130	0.00
29 S	Terphenyl-d14	0.400	0.435	-8.7	125	0.00
30	Benzo(a)anthracene	0.400	0.401	-0.3	115	0.00
31	Chrysene	0.400	0.421	-5.2	119	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.267	33.3#	79	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.399	0.3	111	0.00
34 I	Perylene-d12	0.400	0.400	0.0	108	0.00
35	Benzo(b)fluoranthene	0.400	0.415	-3.7	116	0.00
36	Benzo(k)fluoranthene	0.400	0.435	-8.7	120	0.00
37 C	Benzo(a)pyrene	0.400	0.412	-3.0	115	0.00
38	Dibenzo(a,h)anthracene	0.400	0.409	-2.2	112	0.00
39	Benzo(g,h,i)perylene	0.400	0.414	-3.5	114	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE041219\
Data File : BE099309.D
Acq On : 12 Apr 2019 14:01
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 12 17:55:14 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 12 17:51:18 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				