

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE041219\
 Data File : BE099309.D
 Acq On : 12 Apr 2019 14:01
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 12 17:55:14 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 12 17:51:18 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	123	0.00
2	1,4-Dioxane	0.545	0.568	-4.2	129	0.00
3	n-Nitrosodimethylamine	0.322	0.282	12.4	112	0.00
4 S	2-Fluorophenol	1.063	0.977	8.1	108	0.00
5 S	Phenol-d6	1.461	1.446	1.0	121	0.00
6	bis(2-Chloroethyl)ether	1.310	1.452	-10.8	136	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	124	0.00
8 S	Nitrobenzene-d5	0.275	0.269	2.2	127	0.00
9	Naphthalene	4.351	4.608	-5.9	131	0.00
10	Hexachlorobutadiene	0.283	0.294	-3.9	129	0.00
11 SURR	2-Methylnaphthalene-d10	0.698	0.737	-5.6	133	0.00
12	2-Methylnaphthalene	0.740	0.811	-9.6	138	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	125	0.00
14 S	2,4,6-Tribromophenol	0.184	0.143	22.3	109	0.00
15 S	2-Fluorobiphenyl	1.572	1.707	-8.6	136	0.00
16	Acenaphthylene	1.895	1.825	3.7	125	0.00
17	Acenaphthene	1.118	1.183	-5.8	136	0.00
18	Fluorene	1.615	1.686	-4.4	130	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
20	4-Bromophenyl-phenylether	0.227	0.223	1.8	129	0.00
21	Hexachlorobenzene	0.216	0.214	0.9	129	0.00
22	Pentachlorophenol	0.090	0.089	1.1	146	0.00
23	Phenanthrene	1.035	1.108	-7.1	136	0.00
24	Anthracene	0.976	0.980	-0.4	130	0.00
25 SURR	Fluoranthene-d10	5.280	5.303	-0.4	127	0.00
26	Fluoranthene	1.528	1.578	-3.3	132	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
28	Pyrene	1.056	1.181	-11.8	130	0.00
29 S	Terphenyl-d14	0.730	0.793	-8.6	125	0.00
30	Benzo(a)anthracene	1.259	1.263	-0.3	115	0.00
31	Chrysene	1.235	1.299	-5.2	119	0.00
32	Bis(2-ethylhexyl)phthalate	1.605	1.070	33.3#	79	0.00
33	Indeno(1,2,3-cd)pyrene	1.412	1.408	0.3	111	0.00
34 I	Perylene-d12	1.000	1.000	0.0	108	0.00
35	Benzo(b)fluoranthene	1.194	1.239	-3.8	116	0.00
36	Benzo(k)fluoranthene	1.162	1.262	-8.6	120	0.00
37 C	Benzo(a)pyrene	1.122	1.157	-3.1	115	0.00
38	Dibenzo(a,h)anthracene	1.116	1.142	-2.3	112	0.00
39	Benzo(g,h,i)perylene	1.112	1.150	-3.4	114	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE041219\
Data File : BE099309.D
Acq On : 12 Apr 2019 14:01
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 12 17:55:14 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 12 17:51:18 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		