

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE041219\
 Data File : BE099320.D
 Acq On : 12 Apr 2019 21:01
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 12 21:41:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 12 20:31:12 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	143	0.00
2	1,4-Dioxane	0.400	0.425	-6.2	152	0.00
3	n-Nitrosodimethylamine	0.400	0.424	-6.0	157	0.00
4 S	2-Fluorophenol	0.400	0.428	-7.0	146	0.00
5 S	Phenol-d6	0.400	0.454	-13.5	160	0.00
6	bis(2-Chloroethyl)ether	0.400	0.436	-9.0	156	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	146	0.00
8 S	Nitrobenzene-d5	0.400	0.495	-23.7	189	0.00
9	Naphthalene	0.400	0.426	-6.5	155	0.00
10	Hexachlorobutadiene	0.400	0.428	-7.0	155	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.442	-10.5	164	0.00
12	2-Methylnaphthalene	0.400	0.433	-8.2	159	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	155	0.00
14 S	2,4,6-Tribromophenol	0.400	0.432	-8.0	187	-0.01
15 S	2-Fluorobiphenyl	0.400	0.418	-4.5	162	0.00
16	Acenaphthylene	0.400	0.411	-2.7	165	0.00
17	Acenaphthene	0.400	0.418	-4.5	166	-0.01
18	Fluorene	0.400	0.422	-5.5	163	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	152	0.00
20	4-Bromophenyl-phenylether	0.400	0.425	-6.2	166	0.00
21	Hexachlorobenzene	0.400	0.420	-5.0	162	0.00
22	Pentachlorophenol	0.400	0.543	-35.8#	237	0.00
23	Phenanthrene	0.400	0.429	-7.2	162	0.00
24	Anthracene	0.400	0.421	-5.2	161	0.00
25 SURR	Fluoranthene-d10	0.400	0.407	-1.7	153	0.00
26	Fluoranthene	0.400	0.409	-2.2	155	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	125	0.00
28	Pyrene	0.400	0.485	-21.2	154	0.00
29 S	Terphenyl-d14	0.400	0.470	-17.5	148	0.00
30	Benzo(a)anthracene	0.400	0.417	-4.2	130	0.00
31	Chrysene	0.400	0.441	-10.2	137	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.461	-15.3	149	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.400	0.0	121	0.00
34 I	Perylene-d12	0.400	0.400	0.0	117	0.00
35	Benzo(b)fluoranthene	0.400	0.421	-5.2	127	0.00
36	Benzo(k)fluoranthene	0.400	0.431	-7.7	129	0.00
37 C	Benzo(a)pyrene	0.400	0.417	-4.2	125	0.00
38	Dibenzo(a,h)anthracene	0.400	0.414	-3.5	123	0.00
39	Benzo(g,h,i)perylene	0.400	0.410	-2.5	122	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE041219\
Data File : BE099320.D
Acq On : 12 Apr 2019 21:01
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 12 21:41:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE040219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 12 20:31:12 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				