

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE041519\
 Data File : BE099328.D
 Acq On : 15 Apr 2019 16:29
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 15 19:03:47 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE040219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 02 17:16:42 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	161#	-0.03
2	1,4-Dioxane	0.545	0.581	-6.6	172#	-0.01
3	n-Nitrosodimethylamine	0.322	0.326	-1.2	169#	-0.03
4 S	2-Fluorophenol	1.063	1.081	-1.7	157#	-0.03
5 S	Phenol-d6	1.461	1.589	-8.8	173#	-0.01
6	bis(2-Chloroethyl)ether	1.310	1.391	-6.2	171#	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	163#	-0.01
8 S	Nitrobenzene-d5	0.275	0.327	-18.9	203#	-0.03
9	Naphthalene	4.351	4.728	-8.7	176#	-0.03
10	Hexachlorobutadiene	0.283	0.296	-4.6	170#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.750	-7.4	178#	-0.02
12	2-Methylnaphthalene	0.740	0.788	-6.5	175#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	172#	-0.03
14 S	2,4,6-Tribromophenol	0.184	0.180	2.2	188#	-0.01
15 S	2-Fluorobiphenyl	1.572	1.652	-5.1	181#	-0.02
16	Acenaphthylene	1.895	1.884	0.6	177#	-0.02
17	Acenaphthene	1.118	1.180	-5.5	186#	-0.02
18	Fluorene	1.615	1.677	-3.8	178#	-0.03
19 I	Phenanthrene-d10	1.000	1.000	0.0	166#	-0.03
20	4-Bromophenyl-phenylether	0.227	0.228	-0.4	170#	-0.03
21	Hexachlorobenzene	0.216	0.224	-3.7	174#	-0.03
22	Pentachlorophenol	0.090	0.107	-18.9	225#	-0.03
23	Phenanthrene	1.035	1.123	-8.5	178#	-0.03
24	Anthracene	0.976	0.998	-2.3	170#	-0.01
25 SURR	Fluoranthene-d10	5.280	5.181	1.9	161#	-0.02
26	Fluoranthene	1.528	1.515	0.9	163#	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	134	-0.02
28	Pvrene	1.056	1.253	-18.7	162#	-0.02
29 S	Terphenyl-d14	0.730	0.842	-15.3	156#	-0.02
30	Benzo(a)anthracene	1.259	1.277	-1.4	136	-0.02
31	Chrysene	1.235	1.320	-6.9	142	-0.01
32	Bis(2-ethylhexyl)phthalate	1.605	1.449	9.7	125	-0.02
33	Indeno(1,2,3-cd)pyrene	1.412	1.407	0.4	130	-0.06
34 I	Perylene-d12	1.000	1.000	0.0	125	-0.03
35	Benzo(b)fluoranthene	1.194	1.246	-4.4	135	-0.03
36	Benzo(k)fluoranthene	1.162	1.278	-10.0	141	-0.03
37 C	Benzo(a)pyrene	1.122	1.152	-2.7	132	-0.03
38	Dibenzo(a,h)anthracene	1.116	1.153	-3.3	131	-0.06
39	Benzo(a,h,i)perylene	1.112	1.146	-3.1	132	-0.06

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE041519\
Data File : BE099328.D
Acq On : 15 Apr 2019 16:29
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 15 19:03:47 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE040219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 02 17:16:42 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			