

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE041819\
 Data File : BE099342.D
 Acq On : 18 Apr 2019 18:03
 Operator : JU/SJ
 Sample : K2425-03
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 KY038PS035-190415

Quant Time: Apr 19 07:11:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	3592	0.40	ng	0.00
7) Naphthalene-d8	10.61	136	13252	0.40	ng	0.01
13) Acenaphthene-d10	14.46	164	9727	0.40	ng	0.00
19) Phenanthrene-d10	17.19	188	29528	0.40	ng	0.00
27) Chrysene-d12	21.37	240	34846	0.40	ng	0.00
34) Perylene-d12	23.86	264	37860	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.39	112	1696	0.19	ng	0.00
5) Phenol-d6	6.98	99	1354	0.11	ng	0.00
8) Nitrobenzene-d5	8.96	82	4402	0.43	ng	0.00
11) 2-Methylnaphthalene-d10	12.19	152	9712	0.42	ng	0.00
14) 2,4,6-Tribromophenol	15.94	330	1866	0.58	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	15450	0.40	ng	0.00
25) Fluoranthene-d10	19.21	212	169484	0.48	ng	0.00
29) Terphenyl-d14	19.81	244	33082	0.48	ng	0.00
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.27	149	12678	0.136	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE041819\
Data File : BE099342.D
Acq On : 18 Apr 2019 18:03
Operator : JU/SJ
Sample : K2425-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
KY038PS035-190415

Quant Time: Apr 19 07:11:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

