

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE041819\
 Data File : BE099343.D
 Acq On : 18 Apr 2019 18:42
 Operator : JU/SJ
 Sample : K2425-04
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 KY038PS036-190415

Quant Time: Apr 19 07:12:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	3539	0.40	ng	0.00
7) Naphthalene-d8	10.61	136	12846	0.40	ng	0.01
13) Acenaphthene-d10	14.46	164	8994	0.40	ng	0.00
19) Phenanthrene-d10	17.19	188	29024	0.40	ng	0.00
27) Chrysene-d12	21.35	240	34184	0.40	ng	-0.01
34) Perylene-d12	23.86	264	37882	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.39	112	1476	0.17	ng	0.00
5) Phenol-d6	6.97	99	1151	0.09	ng	0.00
8) Nitrobenzene-d5	8.96	82	3992	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.19	152	9048	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.94	330	1283	0.43	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	14102	0.40	ng	0.00
25) Fluoranthene-d10	19.22	212	153903	0.44	ng	0.00
29) Terphenyl-d14	19.81	244	29687	0.44	ng	0.00
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.26	149	14504	0.159	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE041819\
Data File : BE099343.D
Acq On : 18 Apr 2019 18:42
Operator : JU/SJ
Sample : K2425-04
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
KY038PS036-190415

Quant Time: Apr 19 07:12:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

