

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE041819\
 Data File : BE099337.D
 Acq On : 18 Apr 2019 14:46
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE041819

Quant Time: Apr 19 07:02:03 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.560	0.565	-0.9	106	-0.01
3	n-Nitrosodimethylamine	0.301	0.305	-1.3	125	-0.01
4 S	2-Fluorophenol	0.977	0.988	-1.1	106	0.00
5 S	Phenol-d6	1.395	1.405	-0.7	106	0.00
6	bis(2-Chloroethyl)ether	1.301	1.313	-0.9	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.01
8 S	Nitrobenzene-d5	0.311	0.321	-3.2	110	0.00
9	Naphthalene	4.364	4.460	-2.2	102	0.00
10	Hexachlorobutadiene	0.277	0.284	-2.5	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.706	0.717	-1.6	103	0.01
12	2-Methylnaphthalene	0.743	0.760	-2.3	104	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.133	0.138	-3.8	104	0.00
15 S	2-Fluorobiphenyl	1.572	1.628	-3.6	101	0.00
16	Acenaphthylene	1.808	1.792	0.9	103	0.00
17	Acenaphthene	1.114	1.154	-3.6	104	0.01
18	Fluorene	1.601	1.637	-2.2	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
20	4-Bromophenyl-phenylether	0.222	0.214	3.6	101	0.00
21	Hexachlorobenzene	0.213	0.212	0.5	102	0.00
22	Pentachlorophenol	0.057	0.049	14.0	110	0.00
23	Phenanthrene	1.059	1.075	-1.5	106	0.00
24	Anthracene	0.942	0.941	0.1	106	0.00
25 SURR	Fluoranthene-d10	4.806	4.586	4.6	102	0.00
26	Fluoranthene	1.418	1.366	3.7	103	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pvrene	1.178	1.222	-3.7	102	0.00
29 S	Terphenyl-d14	0.787	0.808	-2.7	101	0.00
30	Benzo(a)anthracene	1.176	1.209	-2.8	108	0.00
31	Chrysene	1.251	1.278	-2.2	99	0.00
32	Bis(2-ethylhexyl)phthalate	1.070	0.968	9.5	106	0.00
33	Indeno(1,2,3-cd)pyrene	1.236	1.292	-4.5	105	0.01
34 I	Perylene-d12	1.000	1.000	0.0	102	0.00
35	Benzo(b)fluoranthene	1.225	1.253	-2.3	105	0.00
36	Benzo(k)fluoranthene	1.221	1.221	0.0	100	0.00
37 C	Benzo(a)pyrene	1.097	1.084	1.2	101	0.00
38	Dibenzo(a,h)anthracene	1.121	1.138	-1.5	104	0.00
39	Benzo(a,h,i)perylene	1.137	1.151	-1.2	103	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE041819\
Data File : BE099337.D
Acq On : 18 Apr 2019 14:46
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE041819

Quant Time: Apr 19 07:02:03 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			