

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE041819\
 Data File : BE099344.D
 Acq On : 18 Apr 2019 19:21
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Apr 19 07:14:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.560	0.648	-15.7	95	0.00
3	n-Nitrosodimethylamine	0.301	0.330	-9.6	106	-0.01
4 S	2-Fluorophenol	0.977	1.244	-27.3#	104	0.00
5 S	Phenol-d6	1.395	1.762	-26.3#	104	0.00
6	bis(2-Chloroethyl)ether	1.301	1.545	-18.8	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
8 S	Nitrobenzene-d5	0.311	0.371	-19.3	104	0.00
9	Naphthalene	4.364	5.082	-16.5	96	0.00
10	Hexachlorobutadiene	0.277	0.327	-18.1	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.706	0.812	-15.0	96	0.00
12	2-Methylnaphthalene	0.743	0.859	-15.6	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.133	0.187	-40.6#	119	0.00
15 S	2-Fluorobiphenyl	1.572	1.804	-14.8	95	0.00
16	Acenaphthylene	1.808	2.063	-14.1	100	0.00
17	Acenaphthene	1.114	1.278	-14.7	97	0.00
18	Fluorene	1.601	1.841	-15.0	97	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4-Bromophenyl-phenylether	0.222	0.247	-11.3	97	0.00
21	Hexachlorobenzene	0.213	0.243	-14.1	97	0.00
22	Pentachlorophenol	0.057	0.072	-26.3#	133	0.00
23	Phenanthrene	1.059	1.188	-12.2	97	0.00
24	Anthracene	0.942	1.053	-11.8	98	0.00
25 SURR	Fluoranthene-d10	4.806	5.459	-13.6	101	0.00
26	Fluoranthene	1.418	1.611	-13.6	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
28	Pvrene	1.178	1.397	-18.6	102	0.00
29 S	Terphenyl-d14	0.787	0.934	-18.7	103	0.00
30	Benzo(a)anthracene	1.176	1.404	-19.4	109	0.00
31	Chrysene	1.251	1.496	-19.6	101	0.00
32	Bis(2-ethylhexyl)phthalate	1.070	1.535	-43.5#	147	0.00
33	Indeno(1,2,3-cd)pyrene	1.236	1.540	-24.6	109	0.00
34 I	Perylene-d12	1.000	1.000	0.0	93	0.00
35	Benzo(b)fluoranthene	1.225	1.438	-17.4	109	0.00
36	Benzo(k)fluoranthene	1.221	1.388	-13.7	103	0.00
37 C	Benzo(a)pyrene	1.097	1.298	-18.3	110	0.00
38	Dibenzo(a,h)anthracene	1.121	1.313	-17.1	109	0.00
39	Benzo(a,h,i)perylene	1.137	1.309	-15.1	106	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE041819\
Data File : BE099344.D
Acq On : 18 Apr 2019 19:21
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 19 07:14:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		