

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE041819\
 Data File : BE099344.D
 Acq On : 18 Apr 2019 19:21
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 19 07:14:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	82	0.00
2	1,4-Dioxane	0.400	0.463	-15.8	95	0.00
3	n-Nitrosodimethylamine	0.400	0.438	-9.5	106	-0.01
4 S	2-Fluorophenol	0.400	0.509	-27.2#	104	0.00
5 S	Phenol-d6	0.400	0.505	-26.2#	104	0.00
6	bis(2-Chloroethyl)ether	0.400	0.475	-18.7	94	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	83	0.00
8 S	Nitrobenzene-d5	0.400	0.477	-19.2	104	0.00
9	Naphthalene	0.400	0.466	-16.5	96	0.00
10	Hexachlorobutadiene	0.400	0.473	-18.2	96	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.460	-15.0	96	0.00
12	2-Methylnaphthalene	0.400	0.463	-15.8	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.400	0.562	-40.5#	119	0.00
15 S	2-Fluorobiphenyl	0.400	0.459	-14.8	95	0.00
16	Acenaphthylene	0.400	0.457	-14.2	100	0.00
17	Acenaphthene	0.400	0.459	-14.8	97	0.00
18	Fluorene	0.400	0.460	-15.0	97	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	88	0.00
20	4-Bromophenyl-phenylether	0.400	0.446	-11.5	97	0.00
21	Hexachlorobenzene	0.400	0.456	-14.0	97	0.00
22	Pentachlorophenol	0.400	0.554	-38.5#	133	0.00
23	Phenanthrene	0.400	0.449	-12.2	97	0.00
24	Anthracene	0.400	0.447	-11.7	98	0.00
25 SURR	Fluoranthene-d10	0.400	0.454	-13.5	101	0.00
26	Fluoranthene	0.400	0.454	-13.5	100	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	88	0.00
28	Pvrene	0.400	0.474	-18.5	102	0.00
29 S	Terphenyl-d14	0.400	0.475	-18.7	103	0.00
30	Benzo(a)anthracene	0.400	0.478	-19.5	109	0.00
31	Chrysene	0.400	0.478	-19.5	101	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.574	-43.5#	147	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.499	-24.7	109	0.00
34 I	Perylene-d12	0.400	0.400	0.0	93	0.00
35	Benzo(b)fluoranthene	0.400	0.470	-17.5	109	0.00
36	Benzo(k)fluoranthene	0.400	0.455	-13.7	103	0.00
37 C	Benzo(a)pyrene	0.400	0.473	-18.2	110	0.00
38	Dibenzo(a,h)anthracene	0.400	0.469	-17.2	109	0.00
39	Benzo(a,h,i)perylene	0.400	0.461	-15.3	106	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE041819\
Data File : BE099344.D
Acq On : 18 Apr 2019 19:21
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 19 07:14:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				