

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE041919\
 Data File : BE099351.D
 Acq On : 19 Apr 2019 16:23
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Apr 19 22:58:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.560	0.659	-17.7	118	0.00
3	n-Nitrosodimethylamine	0.301	0.299	0.7	117	0.00
4 S	2-Fluorophenol	0.977	1.156	-18.3	118	0.00
5 S	Phenol-d6	1.395	1.661	-19.1	119	0.00
6	bis(2-Chloroethyl)ether	1.301	1.513	-16.3	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.311	0.354	-13.8	121	0.00
9	Naphthalene	4.364	5.070	-16.2	116	0.00
10	Hexachlorobutadiene	0.277	0.320	-15.5	114	-0.01
11 SURR	2-Methylnaphthalene-d10	0.706	0.824	-16.7	118	0.00
12	2-Methylnaphthalene	0.743	0.865	-16.4	118	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.133	0.166	-24.8	129	-0.01
15 S	2-Fluorobiphenyl	1.572	1.796	-14.2	115	0.00
16	Acenaphthylene	1.808	1.988	-10.0	117	0.00
17	Acenaphthene	1.114	1.235	-10.9	114	0.00
18	Fluorene	1.601	1.825	-14.0	116	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
20	4-Bromophenyl-phenylether	0.222	0.247	-11.3	114	0.00
21	Hexachlorobenzene	0.213	0.247	-16.0	116	0.00
22	Pentachlorophenol	0.057	0.063	-10.5	138	0.00
23	Phenanthrene	1.059	1.213	-14.5	116	0.00
24	Anthracene	0.942	1.035	-9.9	114	0.00
25 SURR	Fluoranthene-d10	4.806	5.471	-13.8	119	0.00
26	Fluoranthene	1.418	1.628	-14.8	119	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
28	Pyrene	1.178	1.384	-17.5	118	0.00
29 S	Terphenyl-d14	0.787	0.944	-19.9	122	0.00
30	Benzo(a)anthracene	1.176	1.334	-13.4	122	0.00
31	Chrysene	1.251	1.512	-20.9	120	0.00
32	Bis(2-ethylhexyl)phthalate	1.070	1.171	-9.4	132	0.00
33	Indeno(1,2,3-cd)pyrene	1.236	1.546	-25.1#	129	0.00
34 I	Perylene-d12	1.000	1.000	0.0	107	0.00
35	Benzo(b)fluoranthene	1.225	1.423	-16.2	125	0.00
36	Benzo(k)fluoranthene	1.221	1.474	-20.7	127	0.00
37 C	Benzo(a)pyrene	1.097	1.293	-17.9	127	0.00
38	Dibenzo(a,h)anthracene	1.121	1.321	-17.8	127	0.00
39	Benzo(g,h,i)perylene	1.137	1.323	-16.4	124	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE041919\
Data File : BE099351.D
Acq On : 19 Apr 2019 16:23
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 19 22:58:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			