

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE042619\
 Data File : BE099372.D
 Acq On : 26 Apr 2019 11:30
 Operator : JU/SJ
 Sample : PB119200BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB119200BL

Quant Time: Apr 27 05:23:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	3218	0.40	ng	0.00
7) Naphthalene-d8	10.60	136	11842	0.40	ng	0.00
13) Acenaphthene-d10	14.46	164	7946	0.40	ng	0.00
19) Phenanthrene-d10	17.19	188	23919	0.40	ng	0.00
27) Chrysene-d12	21.37	240	28531	0.40	ng	0.00
34) Perylene-d12	23.86	264	31964	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.40	112	4061	0.52	ng	0.01
5) Phenol-d6	6.97	99	5410	0.48	ng	0.00
8) Nitrobenzene-d5	8.97	82	4409	0.48	ng	0.00
11) 2-Methylnaphthalene-d10	12.20	152	8878	0.42	ng	0.00
14) 2,4,6-Tribromophenol	15.94	330	1610	0.61	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	13270	0.42	ng	0.00
25) Fluoranthene-d10	19.22	212	134915	0.47	ng	0.00
29) Terphenyl-d14	19.82	244	23741	0.42	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE042619\
Data File : BE099372.D
Acq On : 26 Apr 2019 11:30
Operator : JU/SJ
Sample : PB119200BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB119200BL

Quant Time: Apr 27 05:23:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

