

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
 Data File : BE099370.D
 Acq On : 26 Apr 2019 10:17
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 27 05:18:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 19 06:51:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	0.00
2	1,4-Dioxane	0.400	0.336	16.0	90	0.00
3	n-Nitrosodimethylamine	0.400	0.351	12.3	111	-0.01
4 S	2-Fluorophenol	0.400	0.469	-17.2	125	0.01
5 S	Phenol-d6	0.400	0.466	-16.5	125	0.00
6	bis(2-Chloroethyl)ether	0.400	0.384	4.0	99	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	0.01
8 S	Nitrobenzene-d5	0.400	0.459	-14.8	131	0.00
9	Naphthalene	0.400	0.428	-7.0	115	0.00
10	Hexachlorobutadiene	0.400	0.435	-8.7	116	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.398	0.5	108	0.01
12	2-Methylnaphthalene	0.400	0.429	-7.2	117	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.400	0.576	-44.0#	152	0.00
15 S	2-Fluorobiphenyl	0.400	0.414	-3.5	107	0.00
16	Acenaphthylene	0.400	0.431	-7.7	118	0.00
17	Acenaphthene	0.400	0.409	-2.2	108	0.01
18	Fluorene	0.400	0.391	2.3	102	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	94	0.00
20	4-Bromophenyl-phenylether	0.400	0.442	-10.5	103	0.00
21	Hexachlorobenzene	0.400	0.441	-10.2	100	0.00
22	Pentachlorophenol	0.400	0.610	-52.5#	166	0.00
23	Phenanthrene	0.400	0.403	-0.8	93	0.00
24	Anthracene	0.400	0.413	-3.2	97	0.00
25 SURR	Fluoranthene-d10	0.400	0.402	-0.5	95	0.00
26	Fluoranthene	0.400	0.397	0.8	94	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
28	Pyrene	0.400	0.407	-1.7	95	0.00
29 S	Terphenyl-d14	0.400	0.408	-2.0	96	0.00
30	Benzo(a)anthracene	0.400	0.453	-13.2	113	0.01
31	Chrysene	0.400	0.406	-1.5	94	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.786	-96.5#	220	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.456	-14.0	109	0.02
34 I	Perylene-d12	0.400	0.400	0.0	108	0.00
35	Benzo(b)fluoranthene	0.400	0.398	0.5	108	0.00
36	Benzo(k)fluoranthene	0.400	0.369	7.8	98	0.00
37 C	Benzo(a)pyrene	0.400	0.409	-2.2	110	0.00
38	Dibenzo(a,h)anthracene	0.400	0.393	1.8	106	0.01
39	Benzo(g,h,i)perylene	0.400	0.384	4.0	103	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE042619\
Data File : BE099370.D
Acq On : 26 Apr 2019 10:17
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Apr 27 05:18:16 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE041819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 19 06:51:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				